

|   |                                   |   |
|---|-----------------------------------|---|
|  | <b>ACCREDITATION<br/>DOCUMENT</b> | <b>F-06/02<br/>Issue Date: 18/08/2020<br/>Rev. No: 09<br/>LAB 158</b> |
|---|-----------------------------------|---|

## **Accreditation No: LAB 158**

### **Awarded to**

**AMSON MEDICAL DEVICES QUALITY CONTROL LABORATORY  
AMSON VACCINES & PHARMA (PVT) Ltd.  
Plot No. 113-Industrial Triangle, Kahuta Road, Islamabad, Pakistan.**

The scope of accreditation is in accordance with the standard specifications outlined in the following page(s) of this document. The accredited scope shall be visible and legible in areas such as customer service, sample-receiving section etc and shall not mislead its users.

The accreditation was first time granted on **12-06-2018** by Pakistan National Accreditation Council.

The laboratory complies with the requirements of **ISO/IEC 17025:2017**.

The accreditation requires regular surveillance, and is valid until **23-07-2024**.

The decision of accreditation made by Pakistan National Accreditation Council implies that the organization has been found to fulfill the requirements for accreditation within the scope.

The organization however, itself is responsible for the results of performed measurements/tests.

**PAKISTAN NATIONAL ACCREDITATION COUNCIL**

15-07-2021

Date

\_\_\_\_\_  
Director General

|   |                                   |   |
|---|-----------------------------------|---|
|  | <b>ACCREDITATION<br/>DOCUMENT</b> | <b>F-06/02</b><br><b>Issue Date: 18/08/2020</b><br><b>Rev. No: 09</b><br><b>LAB 158</b> |
|---|-----------------------------------|---|

### Testing Laboratory.

Accreditation Scope of **QUALITY CONTROL LABORATORY-  
AMSON VACCINES & PHARMA (PVT) Ltd.**  
**Plot No. 113-Industrial Triangle, Kahuta Road, Islamabad,  
Pakistan.**

Permanent laboratory premises

| Materials/<br>Products<br>tested   | Testing field (e.g.<br>environmental testing or<br>mechanical testing) | Types of test/<br>Properties<br>measured | Reference to standardized<br>method (e.g. ISO 14577-<br>1:2003)/ Internal method<br>reference |
|------------------------------------|--|--|---|
| Medical<br>Devices<br>( Syringes ) | Chemical Test  | UV absorbance                            | BP 2021, Vol-5, Page #<br>Appendix XIX G V-A647/<br>UV - Vis.<br>Spectrophotometer            |
| Medical<br>Devices<br>( Syringes ) | Physical Test  | Appearance of<br>Solution                | BP 2021, Vol-5, Page #<br>Appendix XIX G V-A647   |
| Medical<br>Devices<br>( Syringes ) | Chemical Test  | Silicon Oil<br>Estimation                | BP 2021, Vol-5, Page #<br>Appendix XIX G V-A647/<br>ISO 7886-1:1997, Clause:8                 |
| Medical<br>Devices<br>( Syringes ) | Physical Test  | Dead Space                               | ISO 7886-1:2017, Annex C  |

15-07-2021  
Date

\_\_\_\_\_  
Director