

	<b>ACCREDITATION DOCUMENT</b>	<b>F-06/02</b> <b>Issue Date: 10/08/15</b> <b>Rev. No: 07</b> <b>LAB 001</b>
---	-------------------------------	---

**Accreditation No: LAB 001**

**Awarded to**

**NATIONAL PHYSICAL & STANDARDS LABORATORY.  
PLOT NO.16, SECTOR H-9/1, ISLAMABAD, PAKISTAN**

The scope of accreditation is in accordance with the standard specifications outlined in the following page(s) of this document. The accredited scope shall be visible and legible in areas such as customer service, sample-receiving section etc. and shall not mislead its users.

The accreditation was first time granted on **12-02-2004** by Pakistan National Accreditation Council.

The laboratory complies with the requirements of **ISO/IEC 17025:2005**.

The accreditation requires regular surveillance, and is valid until **11-02-2022**.

The decision of accreditation made by Pakistan National Accreditation Council implies that the organization has been found to fulfill the requirements for accreditation within the scope.

The organization however, itself is responsible for the results of performed measurements/tests.

**PAKISTAN NATIONAL ACCREDITATION COUNCIL**

11-11-2019  
Date

\_\_\_\_\_  
Director General



**ACCREDITATION DOCUMENT**

**F-06/02**  
**Issue Date: 10/08/15**  
**Rev. No: 07**  
**LAB 001**

**(A) TESTING LABORATORY**

**Permanent laboratory premises (Testing Scope)**

<b>Chemical Metrology AAS / PCS / Viscosity (Water / Wastewater/ Industrial Effluents/ Solutions/ Liquids /Fluids)</b>			
<b>Materials/ Products Tested</b>	<b>Testing Field (e.g. environmental testing or mechanical testing)</b>	<b>Types of Tests/ properties measured</b>	<b>Reference to Standardized method (e.g. ISO 14577-1:2003)/ Internal method Reference</b>
Water/ Wastewater/ Industrial Effluents	Environmental testing	Measurement of Electrical Conductivity	2510-B, Standard Methods for the Examination of Water and Waste water, 22 <sup>nd</sup> edition, 2012, AWWA/APHA
		Total Dissolved Solids (TDS)	2540-B, Standard Methods for the Examination of Water and Waste water, 22 <sup>nd</sup> edition, 2012, AWWA/APHA
		Total Suspended Solids (TSS)	2540-B, Standard Methods for the Examination of Water and Waste water, 22 <sup>nd</sup> edition, 2012, AWWA/APHA
		Total Hardness	2340-B, Standard Methods for the Examination of Water and Waste water, 22 <sup>nd</sup> edition, 2012, AWWA/APHA
		Alkalinity	2320-B, Standard Methods for the Examination of Water and Waste water, 22 <sup>nd</sup> edition, 2012, AWWA/APHA
		Chloride	4500-C, Standard Methods for the Examination of Water and Waste water, 22 <sup>nd</sup> edition, 2012, AWWA/APHA
		Calcium	3500-Ca, Standard Methods for the Examination of Water and Waste water, 22 <sup>nd</sup> edition, 2012, AWWA/APHA
		Chemical Oxygen Demand (COD)	5220-B, Standard Methods for the Examination of Water and Waste water, 22 <sup>nd</sup> edition, 2012, AWWA/APHA
Water/ Wastewater/ Industrial Effluents/ Solutions/ Liquids	Chemical and Environmental testing	Copper (Cu)	Standard Methods for the Examination of Water and Waste water, 22 <sup>nd</sup> edition, 2012, APHA, AWWA, WEF / Atomic Absorption Spectrometer, A Analyst -100, Perkin Elmer-USA / Direct method (Flame - AAS)
		Iron (Fe)	
		Lead (Pb)	
		Zinc (Zn)	
		Chromium (Cr)	
		Sodium (Na)	
		Potassium (K)	
		Calcium (Ca)	
		Magnesium (Mg)	
Cadmium (Cd)			

11-11-2019  
Date

\_\_\_\_\_  
Director



**ACCREDITATION DOCUMENT**

**F-06/02**  
**Issue Date: 10/08/15**  
**Rev. No: 07**  
**LAB 001**

		Nickel (Ni)	
		Manganese (Mn)	
		Cobalt (Co)	
Liquids / Fluids	Petroleum	Kinematic Viscosity	ASTM D 445-09

**(B) CALIBRATION LABORATORY**

**Permanent laboratory premises**

<b>(i-a) Field of measurement: ELECTRICAL MEASUREMENTS(Measure Mode Scope)</b>				
Measured Quantity	RANGE	Calibration & Measurement Capability Expressed as An Uncertainty ( $\pm$ )	Brief description of measurement & equipment used	
			Equipment used	Method used
<b>DC VOLTAGE</b>	100 mV	0.00085 mV	1. <b>Reference Standard</b> 3458-Agilent, 34401 – Agilent, 8508- A, Fluke, Digital multi-meters High voltage divider DVR 150 2. <b>Unit under test</b> Fluke-5720-a, 9100-Wavetek, Multifunction Calibrator, High voltage source)	EMD/TPP-08 EMD/TPP-10
	1 V	0.000042 V		
	10 V	0.000037 V		
	100 V	0.00056 V		
	1000 V 150kV	0.0056 V 0.5%		
<b>DC CURRENT</b>	100 $\mu$ A	0.0019 $\mu$ A	1. <b>Reference Standard</b> 3458-Agilent, 34401 – Agilent, 8508- A, Fluke, Digital Multi-meters, Clamp meter Radian TX-21 2. <b>Unit under test</b> Fluke-5720-A, 9100-Wavetek Multifunction Calibrator (Source)	EMD/TPP-09 EMD/TPP-10
	1 mA	0.000015 mA		
	10 mA	0.00015 mA		
	100 mA	0.0019 mA		
	3 A	0.00002 A		
	20 A 2000A	0.00024 A 1.5%		
<b>AC VOLTAGE</b>	10 mV, at 50Hz to 300Hz	0.00386 mV	1. <b>Reference Standard</b> 3458-Agilent, 34401 – Agilent, 8508- A,	EMD/TPP-08 EMD/TPP-10
	100 mV, at 50Hz to 300Hz	0.00039 mV		
	1 V, at 50 to 500Hz	0.00021 V		

11-11-2019  
Date

\_\_\_\_\_  
Director



**ACCREDITATION DOCUMENT**

**F-06/02**  
**Issue Date: 10/08/15**  
**Rev. No: 07**  
**LAB 001**

	10 V, at 50 Hz to 10kHz	0.00381 V	Fluke, Digital Multi-meters (Measure) High Voltage Divider 2. <b>Unit Under Test</b> Fluke-5720-A, 9100- Wavetek Multifunction Calibrator (Source), High Voltage Source	
	100 V, at 50 Hz to 10kHz	0.0175 V		
	1000 V, at 50 Hz to 10kHz	0.028 V		
	1kV to 150kV, at 50Hz to 1kHz	0.5%		
<b>AC CURRENT</b>	100 $\mu$ A, at 10 Hz to 1kHz	0.012 $\mu$ A	1 <b>Reference Standard</b> 3458-Agilent, 34401 – Agilent, 8508- A, Fluke, Digital Multi-meters (Measure), Clamp meter 2. <b>Unit Under Test</b> Fluke-5720-a, 9100- Wavetwk Multifunction Calibrator (Source)	EMD/TPP-09 EMD/TPP-10
	10 mA, at 10 Hz to 1kHz	0.0013 mA		
	100 mA, at 10 Hz to 1kHz	0.013 mA		
	3 A, at 10 Hz to 1kHz	0.00013 A		
	20 A, at 10 Hz to 1kHz	0.00634 A		
	2000 A, at 50Hz to 1kHz	1.5%		
<b>RESISTANCE</b>	10 $\Omega$	0.00017 $\Omega$	1 <b>Reference Standard</b> 3458-Agilent, 34401 – Agilent, 8508- A, Fluke, Digital Multi-meters (Measure) 2. <b>Unit Under Test</b> Fluke-5720-A, 9100- Wavetek Multifunction Calibrator (Source)	EMD/TPP-10 EMD/TPP-11 EMD/TPP-12
	100 $\Omega$	0.0014 $\Omega$		
	1 k $\Omega$	0.0000077 k $\Omega$		
	100 k $\Omega$	0.00069 k $\Omega$		
	1 M $\Omega$	0.00001 M $\Omega$		
	100 M $\Omega$	0.014 M $\Omega$		
	2G $\Omega$	0.00026G $\Omega$		
	20G $\Omega$	0.0024G $\Omega$		
<b>CAPACITANCE</b>	10 pF	0.003 pF	1 <b>Reference Standard</b> <b>PM-6306</b> <b>Fluke</b> ,Reference Digital Capacitance Meter (Measure) 2. <b>Unit Under Test</b> 9100-Wavetek, Universal Calibration System (Sources)	EMD/TPP-14
	1000 pF	3.16 pF		
	1 nF	0.001nF		
	1000 nF	2.89 nF		
	1 $\mu$ F	0.0001 $\mu$ F		
	1000 $\mu$ F	0.058 $\mu$ F		
<b>INDUCTANCE</b>	100 $\mu$ H	0.05 $\mu$ H	1. <b>Fixed value working</b> standard Inductors (source) 2. <b>Unit under test</b> Digital inductance/ LCR meter pm-6304 fluke, (measure)	EMD/TPP-15
	1 mH	0.0058 mH		
	10 mH	0.0058 mH		
	100 mH	0.008 mH		
	1 H	0.0001 H		

**(i-b) ELECTRICAL MEASUREMENTS (Source/Generation Scope):**

Measured Quantity	Range	Calibration & measurement capability expressed as an uncertainty ( $\pm$ )	Brief description of measurement & equipment used	
			Equipment used	Method used
<b>DC VOLTAGE</b>	1.018 V	0.2 $\mu$ V	1. <b>Reference Standard</b> Fluke – 732-B, Reference DC Voltage Standard (Source) 2. <b>Unit Under Test</b>	EMD/TPP-10
	10 V	0.4 $\mu$ V		

11-11-2019

Date

Director



**ACCREDITATION DOCUMENT**

**F-06/02**  
**Issue Date: 10/08/15**  
**Rev. No: 07**  
**LAB 001**

			3458-Agilent Reference, 8508- A, Fluke, Digital Multi-meter (Measure)	
	220mV	0.00205 mV	3. <b>Reference Standard</b> Fluke – 5720-A Multifunction Calibrator (Source) 4. <b>Unit Under Test</b> All Types OF Digital / Analog Multi-meter 3458-AGILENT, 8508-A, Fluke, 45-Fluke, ETC (Measure)	EMD/TPP-10 EMD/TPP-08
	2.2 V	0.0000117 V		
	11 V	0.000041 V		
	22 V	0.000081 V		
	220 V	0.00114 V		
1100 V	0.00755 V			
<b>DC CURRENT</b>	220 $\mu$ A	0.148 $\mu$ A	1. Reference Standard Reference <b>Standard, Fluke – 5720-A</b> Multifunction Calibrator-9100 Source) 2. <b>Unit Under Test</b> All Types OF Digital / Analog Multi-meters 3458-Agilent, 8508- A, Fluke, 45-Fluke, Clamp meter,etc (Measure)	EMD/TPP-10 EMD/TPP-09
	2.2 mA	0.000084 mA		
	22 mA	0.000811 mA		
	220 mA	0.0106 mA		
	2.2 A	0.000188 A		
	20 A, 1000 A (via current coil)	0.00408 A 0.00444 A		
<b>AC VOLTAGE</b>	2.2mV, at 10 Hz to 100kHz	0.02594 mV	1. Reference Standard Reference <b>Standard, Fluke – 5720-A</b> Multifunction Calibrator (Source) 2. <b>Unit Under Test</b> All Types OF Digital / Analog Multi-meters 3458-Agilent, 8508- A, Fluke, 45-Fluke, etc. (Measure)	EMD/TPP-10 EMD/TPP-08
	22 mV, at 10 Hz to 100kHz	0.0794 mV		
	220 mV, at 10 Hz to 100kHz	0.639 mV		
	2.2 V, at 10 Hz to 100kHz	0.00404 V		
	22 V, at 10 Hz to 100kHz	0.0362V		
	220 V, at 10 Hz to 100kHz	0.0568V		
	750 V, at 40 Hz to 10 kHz	0.461V		
	1100 V, at 40 Hz to 10 kHz	0.671V		
<b>AC CURRENT</b>	220 $\mu$ A, at 10 Hz to 10 kHz	0.307 $\mu$ A	1 <b>Reference Standard</b> Fluke – 5720-A Multifunction Calibrator - 9100 (Source) 2. <b>Unit Under Test</b> All Types of Digital / Analog Multi-meter i.e 3458-Agilent, 8508-A, Fluke, 45-Fluke, clamp meters etc (Measure)	EMD/TPP-10 EMD/TPP-09
	2.2 mA, at 10 Hz to 10 kHz	0.00307 mA		
	22 mA, at 10 Hz to 10 kHz	0.0292 mA		
	220 mA, at 10 Hz to 10 kHz	0.252 mA		
	2.2 A, at 20 Hz to 10 kHz	0.01556 A		
	20 A, at 40 Hz to 10 kHz 1000 A, at 40Hz to 1kHz (via current coil)	0.04035 A		
<b>RESISTANCE</b>	1 $\Omega$	0.000095 $\Omega$	1 <b>Reference Standard</b> Fluke – 5720-A Multifunction Calibrator (Source) 2. <b>Unit Under Test</b> All Types OF Digital / Analog Multi-meter i.e 3458-Agilent, 8508-A, Fluke, 45-Fluke, etc (Measure)	EMD/TPP-10 EMD/TPP-11
	10 $\Omega$	0.000023 $\Omega$		
	100 $\Omega$	0.00001 $\Omega$		
	1 k $\Omega$	0.000009 k $\Omega$		
	10 k $\Omega$	0.000009 k $\Omega$		
	100 k $\Omega$	0.00001 $\Omega$		
	1 M $\Omega$	0.00002 M $\Omega$		
	10 M $\Omega$	0.00005 M $\Omega$		
100 M $\Omega$	0.0001M $\Omega$			

11-11-2019

Date

Director



**ACCREDITATION DOCUMENT**

**F-06/02**  
**Issue Date: 10/08/15**  
**Rev. No: 07**  
**LAB 001**

<b>RESISTANCE</b>	0.0001 Ω	60 μΩ	1. Hi-Accuracy Working Standard 4-Terminal Standard Resistors <b>20-E/D to 28-E/D, H&amp;B Germany (Source)</b> 2. <b>Unit Under Test</b> 34420-Agilent Digital μΩ Meter (Measure)	EMD/TPP-12
	10 kΩ	0.008 Ω		
<b>CAPACITANCE</b>	0.1 μF	0.0005 μF	1. Fixed Value Working Standard Capacitors (Source) 2. Unit Under Test Digital Capacitance/ LCR Meter <b>PM-6304 Fluke, (Measure)</b>	EMD/TPP-13
	1 μF	0.0005 μF		
	4 μF	0.05 % + 3 pF	1. Reference Standard <b>9100-Wavetek, Universal</b> Calibration system (Source) 2. Unit Under Test <b>Digital Capacitance Meter LCR Meter,</b> (Measure)	EMD/TPP-13
	40 μF	0.05 %		
	400 μF	0.05 %		
	4 mF	0.05 %		
40 mF	0.1 %			
<b>DC POWER</b>	1W - 500kW	0.1W to 0.155 kW	1. Reference Standard <b>9100-Wavetek, Universal</b> Calibration system (Source) 2. Unit Under Test <b>Power Analyzer / Wattmeter</b> (Measure)	EMD/TPP-14
<b>AC POWER</b>	1W - 500kW, at 50Hz – 1kHz	0.1W to 0.155 kW	1. Reference Standard Universal Calibration System <b>9100-Wavetek (Source)</b> 2. Unit Under Test <b>Power Analyzer</b> (Measure)	EMD/TPP-14
<b>INDUCTANCE</b>	100 μH	0.008 μH	1. Fixed Value Working standard <b>Inductors (Source)</b> 2. Unit Under Test Digital Inductance/ LCR Meter <b>PM-6304 Fluke, (Measure)</b>	EMD/TPP-15
	1 mH	0.006 mH		
	10 mH	0.006 mH		
	100 mH	0.006 mH		
	1 H	0.006 H		

**(ii) Field of Measurement: Time & Frequency Measurements**

11-11-2019

Date

Director



## ACCREDITATION DOCUMENT

**F-06/02**  
**Issue Date: 10/08/15**  
**Rev. No: 07**  
**LAB 001**

Measured Quantity	Range	Calibration & Measurement Capability (CMC) expressed as an uncertainty (Expanded) (±)	Brief description of measurement and equipment used	
			Equipment Used	Method Used
<b>Frequency (Source)</b>	10-MHz	2.56E-09 MHZ	<b>Reference Standards:</b> a) Cesium Frequency Standard 5071A b) Agilent Universal Frequency Counter/Timer 53230A(Measure)  <b>Unit Under Test:</b> Precision Test Systems RfFS10,Rb.Frequency Standard (Source)	Direct Frequency Comparison with Frequency counter (T&FMD/TPP -03)
<b>Frequency (Measure)</b>	10 Hz to 500 Hz @ 0dBm	2.91E-03	<b>Reference Standards:</b> c) Cesium Frequency Standard 571A a) Marconi Instrument, AM/FM signal Generator 2024 (Source) b) SMF100 Signal Generator  <b>Unit Under Test:</b> a) Agilent Universal Frequency Counter/Timer 53230A(Measure) b) Microwave Frequency Counter 53152A	Direct Frequency Comparison (T&FMD/TPP -01)
	500 Hz to 10kHz @ 0dBm	2.91E-03 to 1.15E-05		
	10 KHz to 500 kHz @ 0dBm	1.15E-05		
	500 kHz to 1 MHz @ 0dBm	1.15 -05 to 1.33E-08		
	1 MHz to 100 MHz @ 0dBm	1.33E-08 to 1.02E-07		
	100 MHz to 500 MHz @ 0dBm	1.02E-07 to 4.28E-07		
	500 MHz to 1 GHz @ 0dBm	4.28E-07 to 1.73E-10		
	1 GHz to 10 GHz @ 0dBm	1.73E-10 to 1.42E-09		
	10 GHz to 20 GHz @ 0dBm	1.42E-09 to 1.98E-09		
	20 GHz to 30 GHz @ 0dBm	1.98E-09 to 3.04E-09		
	30 GHz to 40 GHz @ 0dBm	3.04E-09 to 1.38E-07		
Time base 10 MHz	4.90E-07			
<b>Frequency (Measure)</b>	10 Hz to 500 Hz @ 0dBm	1.15E-02	<b>Reference Standards:</b> d) Cesium Frequency Standard 5071A c) Agilent Universal Frequency Counter/Timer 53230A(Measure) d) Microwave Frequency Counter 53152A  <b>Unit Under Test:</b> Marconi Instrument, AM/FM signal Generator 2024 (Source) a) SMF100 Signal Generator	Direct Frequency Comparison with Frequency Counter (T&FMD/TPP -04)
	500 Hz to 10kHz @ 0dBm	1.15E-02 to 1.15E-05		
	10 KHz to 500 kHz @ 0dBm	1.15E-05 to 8.00E-05		
	500 kHz to 1 MHz @ 0dBm	8.00E-05 to 1.16E-08		
	1 MHz to 100 MHz @ 0dBm	1.16E-08 to 4.13E-07		
	100 MHz to 500 MHz @ 0dBm	4.13E-07 to 1.73E-08		
	500 MHz to 1 GHz @ 0dBm	.73E-08 to 2.91E-10		
	1 GHz to 10 GHz @ 0dBm	2.91E-10 to 1.42E-09		
	10 GHz to 20 GHz @ 0dBm	1.42E-09 to 2.61E-09		
	20 GHz to 30 GHz @ 0dBm	2.61E-09 to 2.42E-09		
	30 GHz to 40 GHz @ 0dBm	2.42E-09 to 2.39E-07		
Time base 10 MHz	1.65E-08			
<b>Time (Measure)</b>	01s to 2-hrs.	2.60E-01s	<b>Reference Standards:</b> e) Cesium Frequency Standard 5071A	Direct Comparison Method (T&FMD/TPP -02)

11-11-2019

Date

Director

	<b>ACCREDITATION DOCUMENT</b>	<b>F-06/02</b> <b>Issue Date: 10/08/15</b> <b>Rev. No: 07</b> <b>LAB 001</b>
---	-------------------------------	---

			a) Precision Test Systems GPS10RBN <b>Unit Under Test:</b> a) Casio, HS-60W-IDF, Q&Q Digital Stopwatch.	
--	--	--	--	--

**(iii) Field of measurement: Pressure Metrology**

MEASURED QUANTITY	RANGE	Calibration & Measurement Capability Expressed As An Uncertainty ( $\pm$ )	BRIEF DESCRIPTION OF MEASUREMENT & EQUIPMENT USED	
			EQUIPMENT USED	METHOD USED
Hydraulic Pressure (0.1 – 1100) bar	(0.1 – 35) bar	0.060 bar	Hydraulic Dead Weight Tester (P- 01) is used to calibrate Dead Weight Testers, Hydraulic Gauges and transducers. Its measurement is traceable via NML, SIRIM.	Direct Comparison
	(36 – 500) bar	0.220 bar		
	(501 – 1100) bar	0.276 bar		
Hydraulic Pressure (10 – 18000) psi	(10 – 1000) psi	1.35 psi	Hydraulic Dead Weight Tester (P- 09) is used to calibrate Dead Weight Testers, Hydraulic Gauges and transducers. Its measurement is traceable via NML, SIRIM.	Direct Comparison
	(1001 – 5000 ) psi	2.29 psi		
	(5001 – 10000) psi	3.70 psi		
	(10001 – 18000) psi	4.50 psi		
Pneumatic Pressure (-1 – 20) bar	(-1 – 0) bar	0.007 bar	Pneumatic Pressure Calibrator (P – 10) is used to calibrate Pneumatic/ vacuum gauges, pressure calibrator, manometers and transducers. . Its measurement is traceable via NVLAP, USA.	Direct Comparison
	(0– 10) bar	0.062 bar		
	(11 – 20) bar	0.039 bar		
Atmospheric Pressure (0.70 – 1.10) bar	(0.70 – 1.10) bar	0.00016 bar	RPM4 Reference Barometer (P-12) is used to calibrate Barometers and Manometers. Its measurement is traceable via aZLa Fluke USA.	Direct Comparison

**(iv) Field of measurement: CONDUCTIVITY MEASUREMENTS**

Measured quantity	Range	Calibration & Measurement Capability (CMC) expressed as an uncertainty( $\pm$ )	Brief description of measurement and equipment used
-------------------	-------	---	---

11-11-2019  
Date

\_\_\_\_\_  
Director

 Pakistan National Accreditation Council	<b>ACCREDITATION DOCUMENT</b>	<b>F-06/02</b> <b>Issue Date: 10/08/15</b> <b>Rev. No: 07</b> <b>LAB 001</b>
--	-------------------------------	---

Calibration of Conductivity Meter	1410 $\mu\text{mhos/cm}$	+5.2 $\mu\text{mhos/cm}$	Calibration of conductivity meters by using KCl standard solution.
-----------------------------------	--------------------------	--------------------------	--

**(v) Field of measurement: pH MEASUREMENTS**

Measured quantity	Range	Calibration & Measurement Capability (CMC) expressed as an uncertainty( $\pm$ )	Brief description of measurement and equipment used
pH	0-14 pH	0.02	Standard Buffer solutions / Two point calibration method

**(vi) : Field of measurement: VISCOSITY MEASUREMENTS**

Measured quantity	Range	Calibration & Measurement Capability (CMC) expressed as an uncertainty ( $\pm$ )	Brief description of measurement and equipment used
UBBELOHDE Viscometer Calibration	(0.003 to 500) $\text{mm}^2 \text{s}^{-2}$	1.05 %	ASTM D 2162-06 Viscosity Standard oils

**(vii) Field of measurement: TEMPERATURE MEASUREMENTS**

MEASURED QUANTITY	RANGE	Calibration & Measurement Capability Expressed As An Uncertainty ( $\pm$ )	BRIEF DESCRIPTION OF MEASUREMENT AND EQUIPMENT USED	
			EQUIPMENT USED	METHOD USED
TEMPERATURE (SOURCES)  Heat Sources	-80.00 °C	0.20°C	<b>1. REFERENCE STANDARDS:-</b> (Black Stack Thermometer FLUKE), (Working Standard PRT FLUKE), (R-Type Thermocouple 5649 FLUKE) (Zero Point Dry Well FLUKE) (Temperature/ Pressure Calibrator 512B )  <b>2. UNIT UNDER TEST:-</b> All types of Heat Sources	TPP-33
	0.00 °C	0.20°C		
	100 °C	1.18°C		
	200 °C	1.14°C		
	299.3 °C	1.13°C		
	349.2 °C	1.14°C		
	399.1 °C	1.14°C		
	800.0 °C	2.00°C		
LIQUID-IN-GLASS THERMOMETER  TEMPERATURE (MEASURE)	-40 °C	0.60 °C	<b>1. REFERENCE STANDARDS:-</b> <b>a)</b> Ultra Low Temperature Bath (Heart Scientific 7380) <b>b)</b> Liquid In Glass Thermometer (ASTM) <b>c)</b> High Precision Thermostatic Bath (Model: PROLABO) <b>d)</b> Oil Bath (HO-21S) <b>e)</b> Working Standard PRT (5628) <b>f)</b> Black Stack Digital Readout (1560) <b>2. UNIT UNDER TEST:-</b> Liquid –In- Glass Thermometers	TPP-25
	0 °C	0.12 °C		
	10 °C	0.12 °C		
	30 °C	0.12 °C		
	50 °C	0.12 °C		
	80 °C	0.20 °C		
	100 °C	1.20 °C		
	150 °C	1.20 °C		
THERMO-HYGROMETER TEMPERATURE /HUMIDITY (SOURCES)	20 %RH	1.38 %RH	<b>1. REFERENCE USED:-</b> <b>a)</b> RH Generator (Humi Lab) <b>b)</b> NESLAB RTE Bath/Circulator (USA) <b>c)</b> Humidity and Temperature Probe (HMP75) (Viasala)  <b>2. UNIT UNDER TEST:-</b> All types of Thermo hygrometers	TPP-32
	30 %RH	1.38 %RH		
	40 %RH	1.38 %RH		
	50 %RH	1.74 %RH		
	60 %RH	1.74 %RH		
	70 %RH	1.86%RH		
	80 %RH	1.86%RH		
	18 °C	0.26 °C		
	20 °C	0.26 °C		
	22 °C	0.26 °C		
	24 °C	0.26 °C		
	26 °C	0.26 °C		
TEMPERATURE (°C) (MEASURE)	-80 °C	0.02 °C	<b>1. Reference Used:-</b> <b>a)</b> Working Standard PRT (5628)	TPP-31
	0 °C	0.06 °C		

11-11-2019  
Date

\_\_\_\_\_  
Director



**ACCREDITATION DOCUMENT**

**F-06/02**  
**Issue Date: 10/08/15**  
**Rev. No: 07**  
**LAB 001**

<b>PLATINUM RESISTANCE THERMOMETER (PRT)</b>	100 °C	0.20 °C	<b>b)</b> Black Stack Digital Readout (1560) <b>c)</b> Dry Block Calibrator ISOTECH GIMNI 700LRI <b>d)</b> Ultra Low Temperature Bath (7380) <b>2. Unit Under Test:-</b> All types of PRT/RTD	
	200 °C	0.22 °C		
	300 °C	0.26 °C		
	400 °C	0.31 °C		
	500 °C	0.40 °C		
	600 °C	0.50 °C		
<b>TEMPERATURE (SOURCES) THERMOCOUPLE</b>	-40 °C	0.2°C	<b>1. Reference Used:-</b> <b>a)</b> Dry Block Calibrator (Model: Gemini700LRI) Ultra Low Temperature Bath (Model:7380 USA), E-Type Thermocouple (Serial#:2H19 Japan) <b>b)</b> Black Stack Digital Readout (Serial#: A8B906, Fluke USA) R-Type Thermocouple (Serial #: 2J13 ,Japan) Furnace (9112B,Fluke USA) <b>c)</b> Temperature/Pressure Calibrator (Model#:525BFluke USA), Working Standard PRT (Model #: 5628) <b>d)</b> Digital Thermometer (YEW 2575) Digital Thermometer (YEW 2572), R-Type Thermocouple Serial#:4996), Muffle Furnace (Model#:KE-6HK1200-3) <b>e)</b> Zero Point Dry Well (Model#:9101) <b>f)</b> S-Type Thermocouple (Model#:5650) <b>2. Unit Under Test:-</b> <b>a)</b> S-Type Thermocouple <b>b)</b> R-Type Thermocouple <b>c)</b> K-Type Thermocouple	TPP-26
	0 °C	0.1°C		
	200 °C	0.2°C		
	400 °C	0.4°C		
	600 °C	0.6°C		
	800 °C	1.5°C		
	1000 °C	1.8°C		
1100 °C	2.0°C			

**CALIBRATION LABORATORY**

**Onsite Accreditation Scope of National Physical & Standards Laboratory (NPSL), Islamabad, Pakistan.**

**Mobile laboratory (Onsite Calibration)**

<b>Field of measurement: i) TEMPERATURE MEASUREMENTS</b>			
<b>Measured quantity</b>	<b>Range</b>	<b>Calibration &amp; Measurement Capability (CMC) expressed as an uncertainty (±)</b>	<b>Brief description of measurement and equipment used</b>
<b>Calibration of Thermocouples</b>	Ambient to 650 °C	1 °C to 1.5 °C	Dry block Temperature Calibrator, Digital Readout, Reference Thermometers, Zero point Dry Well.
<b>Calibration of PRT/SPRT</b>	Ambient to 650 °C	0.15 °C to 0.5 °C	Reference Thermometers, Digital Readout, Zero point Dry

11-11-2019  
Date

\_\_\_\_\_  
Director

	<b>ACCREDITATION DOCUMENT</b>	<b>F-06/02</b> <b>Issue Date: 10/08/15</b> <b>Rev. No: 07</b> <b>LAB 001</b>
---	-------------------------------	---

			Well, Dry block Temperature Calibrator.
<b>Calibration of Low Temperature Bath/ Heat Sources (Oven, Incubator, Dry Block, Furnace)</b>	-80°C to 1400°C	± 0.1 °C to ± 0.5 °C	PRT, S & R Type Thermometer.
<b>Calibration of Thermo Hygrometers</b>	20%RH to 80%RH	1.4% RH to 1.9 %RH	RH Generator Humi-Lab. Humidity and Temperature Probe (VAISALA).

**Field of measurement: ii) ELECTRICAL MEASUREMENTS**

Measured Quantity	RANGE	Calibration & Measurement Capability Expressed as An Uncertainty (±)	Brief description of measurement & equipment used
			Equipment used
<b>DC VOLTAGE</b>	100 mV – 1000V	0.00085 mV - 0.0056 V	<ul style="list-style-type: none"> <li>• <b>Reference Standard</b> Digital multi-meters High voltage divider DVR 150</li> <li>• <b>Unit under test</b> Multifunction Calibrator, High voltage source</li> </ul>
	1kV - 150kV	0.5%	
<b>AC VOLTAGE</b>	100 mV – 1000V at 50Hz to 10kHz	0.0039 mV - 0.028 V	-----do-----
	1kV - 150kV at 50Hz to 1kHz	0.5%	
<b>DC CURRENT</b>	1 mA – 20 A	0.00015 mA – 0.00025A	<ul style="list-style-type: none"> <li>• <b>Reference Standard</b> 3458-Agilent, 34401 – Agilent, 8508- A, Fluke, Digital Multi-meters, Clamp meter Radian TX-21</li> <li>• <b>Unit under test</b> Multifunction Calibrator High Current Source</li> </ul>
	20 A - 2000A	1.5%	
<b>AC CURRENT</b>	1mA to 20A, at 10 Hz to 1kHz	0.0013 mA	-----do-----
	20A - 2000A, at 50Hz to 10kHz	1.5%	
<b>RESISTANCE</b>	1Ω - 2GΩ	0.00017 Ω - 0.00026GΩ	<ul style="list-style-type: none"> <li>• <b>Reference Standard,</b> Digital Multi-meters</li> <li>• <b>Unit Under Test</b> Resistance Calibrator Fixed / Decade Resistance Box</li> </ul>
	2GΩ - 20GΩ	0.0026GΩ	
<b>CAPACITANCE</b>	1nF – 1000µF	0.001nF – 0.058µF	<ul style="list-style-type: none"> <li>• Reference Standard <b>PM-6306 Fluke,</b></li> <li>• <b>Unit Under Test</b> Decade / Fixed Capacitance Standards</li> </ul>
<b>INDUCTANCE</b>	100 µH – 1H	0.05 µH – 0.0061H	<ul style="list-style-type: none"> <li>• Reference Standard, <b>PM-6306 Fluke,</b></li> <li>• <b>Unit under test</b> Digital / Fixed / Decade Inductance</li> </ul>

11-11-2019  
Date

\_\_\_\_\_  
Director