

	<b>ACCREDITATION DOCUMENT</b>	<b>F-06/02 Issue Date: 18/08/2020 Rev. No: 09 LAB 001</b>
---	-----------------------------------	---

## **Accreditation No: LAB 001**

### **Awarded to**

### **NATIONAL METROLOGY INSTITUTE OF PAKISTAN(NMIP) PLOT NO. 16, SECTOR H-9/1, ISLAMABAD, PAKISTAN.**

The scope of accreditation is in accordance with the standard specifications outlined in the following page(s) of this document. The accredited scope shall be visible and legible in areas such as customer service, sample-receiving section etc and shall not mislead its users.

The accreditation was first time granted on **12-02-2004** by Pakistan National Accreditation Council.

The laboratory complies with the requirements of **ISO/IEC 17025:2017**.

The accreditation requires regular surveillance, and is valid until **11-08-2025**.

The decision of accreditation made by Pakistan National Accreditation Council implies that the organization has been found to fulfill the requirements for accreditation within the scope.

The organization however, itself is responsible for the results of performed measurements/tests.

### **PAKISTAN NATIONAL ACCREDITATION COUNCIL**

19-05-2025  
Date

SD  
Director General

 <p><b>PNAC</b> Pakistan National Accreditation Council</p>	<b>ACCREDITATION DOCUMENT</b>	<b>F-06/02</b> <b>Issue Date: 18/08/2020</b> <b>Rev. No: 09</b> <b>LAB 001</b>
--	-----------------------------------	---

## Testing Laboratory

Accreditation Scope of **NATIONAL METROLOGY INSTITUTE OF PAKISTAN**  
Plot No. 16, Sector H-9/1, Islamabad, Pakistan

Permanent laboratory premises

Materials/ Products tested	Testing field (e.g. environmental testing or mechanical testing)	Types of tests / Properties measured	Reference to standardized method (e.g. ISO 14577-1:2003)/ Internal method reference		
<b>CHEMICAL METROLOGY</b>					
<b>Water/ Wastewater/ Industrial Effluents</b>	Environmental testing	Electrical Conductivity	2510-B,		
		Total Dissolved Solids (TDS)	2540-C		
		Total Suspended Solids (TSS)	2540-B		
		Total Hardness	2340-C		
		Alkalinity	2320-B		
		Chloride	4500-Cl <sup>-</sup>		
		Calcium	3500-Ca		
		Magnesium	3500-Mg(B)		
		Chemical Oxygen Demand (COD)	5220-B,		
		pH	4500-H <sup>+</sup>	Standard Methods for the Examination of Water and Wastewater, 24th Edition, 2023. Lipps WC, Braun-Howland EB, and Baxter TE (Eds). American Public Health Association (APHA), American Water Works Association (AWWA), WEF, USA.	
<b>Water/ Wastewater/ Industrial Effluents/ Solutions/ Liquids</b>	Chemical and Environmental testing	Copper (Cu)	Standard Methods for the Examination of Water and Wastewater, 24th Edition, 2023. Lipps WC, Braun-Howland EB, and Baxter TE (Eds). American Public Health Association (APHA), American Water Works Association (AWWA), WEF, USA.		
		Iron (Fe)			
		Lead (Pb)			
		Zinc (Zn)			
		Chromium (Cr)			
		Sodium (Na)			Atomic Absorption Spectrometer, AAnalyst -100, Perkin Elmer-USA Direct method (Flame - AAS)
		Potassium (K)			
		Calcium (Ca)			
		Magnesium (Mg)			
		Cadmium (Cd)			
		Nickel (Ni)			
		Manganese (Mn)			
		Cobalt (Co)			

19-05-2025  
Date

\_\_\_\_\_  
Sd  
Director



## ACCREDITATION DOCUMENT

**F-06/02**  
**Issue Date: 18/08/2020**  
**Rev. No: 09**  
**LAB 001**

### Calibration Laboratory

Accreditation Scope of **NATIONAL METROLOGY INSTITUTE OF PAKISTAN**

Plot No. 16, Sector H-9/1, Islamabad, Pakistan

Permanent laboratory premises

#### Field of measurement: MASS METROLOGY

Measured quantity	Range	*Expanded Uncertainty ( $\pm$ )	Technique, Reference Standard, Equipment
<b>Mass</b>	1 mg - 10 mg	0.0011 - 0.0013 mg	<b>i)</b> E <sub>2</sub> , F <sub>1</sub> and F <sub>2</sub> class standard masses from 1 mg to 10 kg <b>ii)</b> Mass Comparator CCE6 having readability 0.1 $\mu$ g and capacity of 6.1 g <b>iii)</b> Mass Comparator CC111 having readability 1 $\mu$ g and capacity of 111 g <b>iv)</b> Mass Comparator AX1006 having readability 1 $\mu$ g and capacity of 1000 g <b>v)</b> Mass Comparator CC10000U-L having readability 10 $\mu$ g and capacity of 10 kg <b>vi)</b> Top loading balance CC64K <b>vii)</b> Top loading Balance LP1200 <u>Unit Under Test:</u> Standard Masses class E2 and below <u>Method Used:</u> MMD/TPP-21/01, 02, 03, 05, 06, 07
	20 mg - 500 mg	0.0014 - 0.0044 mg	
	1 g - 5 g	0.0044 - 0.0082 mg	
	10 g - 200 g	0.0090 - 0.048 mg	
	500 g - 1000 g	0.16 - 0.29 mg	
	2000 g - 5000 g	0.52 - 1.40 mg	
	10000 g	2.80 mg	

#### Field of measurement: VOLUME METROLOGY

Measured quantity	Range	*Expanded Uncertainty ( $\pm$ )	Technique, Reference Standard, Equipment
<b>Volume</b>	200 $\mu$ L - 5000 $\mu$ L	0.60 $\mu$ L - 2.90 $\mu$ L	<u>Unit Under Test:</u> Micropipette <u>Equipment Used:</u> M-39 ME414S, Sartorius, Germany M-58 CP8200, Sartorius, Germany M-28 CC64K, Sartorius, Germany M-27 LP1200S, Sartorius, Germany <u>Method Used:</u> ASTM E542-01 MMD/Vol-28/TPP-28
	1000 $\mu$ L - 5000 $\mu$ L	2.90 $\mu$ L - 5.00 $\mu$ L	
	1 mL - 1000 mL	0.010 mL - 3.50 mL	<u>Unit Under Test:</u> Glassware
	1 L - 5 L	0.0040 L - 0.030 L	
	20 mL - 2000 mL	0.12 mL - 5.78 mL	<u>Unit Under Test:</u> Metallic
	2 L - 5 L	0.006 L - 0.017 L	
	5 L - 10 L	0.017 L - 0.029 L	
10 L - 20 L	0.029 L - 0.050 L		

#### Field of measurement: DIMENSION METROLOGY

Measured quantity	Range	*Expanded Uncertainty ( $\pm$ )	Technique, Reference Standard, Equipment
<b>Length</b>	0-300 mm / 0-30 cm / 1 Foot	0.041 - 0.071 mm	<u>Unit Under Test:</u> Foot Scales using Length Comparator <u>Method Used:</u>

19-05-2025  
Date

\_\_\_\_\_  
Sd  
Director



**ACCREDITATION  
DOCUMENT**

**F-06/02**  
**Issue Date: 18/08/2020**  
**Rev. No: 09**  
**LAB 001**

<b>Length</b>			DMD/TP-1
	0-1000 mm (1 Meter)	0.041 - 0.062 mm	<b>Unit Under Test:</b> Meter Scales using Length Comparator <b>Method Used:</b> DMD/TP-1
	0.01 - 50 m	0.053 - 0.19 mm	<b>Unit Under Test:</b> Measuring Tapes using Length Comparator <b>Method Used:</b> DMD/TP-1
	1000 mm	0.0015 - 0.0050 mm	<b>Unit Under Test:</b> End Standard Meter Gauge Block using End to End Comparator <b>Method Used:</b> DMD/TP-2
	0.03 – 1 mm	0.20 - 3.00 $\mu$ m	<b>Unit Under Test:</b> Feeler Gauge using UMM <sub>2</sub> <b>Method Used:</b> DMD/TP-3
	0.5 - 100 mm	0.11 – 15.00 $\mu$ m	<b>Unit Under Test:</b> Gauge Blocks using UMM <sub>2</sub> <b>Method Used:</b> DMD/TP-3
	101 - 300 mm	0.13 - 4.8 $\mu$ m	<b>Unit Under Test:</b> Gauge Blocks using UMM <sub>2</sub> <b>Method Used:</b> DMD/TP-3
	0 - 300 mm	0.10 - 25.00 $\mu$ m	<b>Unit Under Test:</b> Precision Glass / Scale / Slide using UMM <sub>1</sub> <b>Technique Used:</b> Direct Measurements
	0 - 80 mm	0.72 – 14.00 $\mu$ m	<b>Unit Under Test:</b> Dial Indicator / Gauge using UMM <sub>1</sub> <b>Method Used:</b> DMD/TP-7
	0 - 300 mm	0.0058 - 0.088 mm	<b>Unit Under Test:</b> Digital Vernier /Dial Caliper using Gauge Blocks <b>Method Used:</b> DMD/TP-4
	0 - 25.4 mm	0.00060 - 0.0090 mm	<b>Unit Under Test:</b> Micrometer using Gauge Blocks <b>Method Used:</b> DMD/TP-5
	0 - 25 mm	0.00060 - 0.0067 mm	<b>Unit Under Test:</b> Dial Gauge Calibrator/ Dial Testing Machine / Tesa Tronic Amplifier using Gauge Blocks <b>Method Used:</b> Direct Measurements
	<b>Diameter External</b>	0.5 - 100 mm	0.18 - 10.00 $\mu$ m
<b>Diameter Internal</b>	5 - 100 mm	0.74 - 16.00 $\mu$ m	<b>Unit Under Test:</b>

19-05-2025  
Date

Sd \_\_\_\_\_  
Director



## ACCREDITATION DOCUMENT

**F-06/02**  
**Issue Date: 18/08/2020**  
**Rev. No: 09**  
**LAB 001**

			Ring Gauges using UMM <sub>1</sub> <b>Method Used:</b> DMD/TP-8		
<b>Field of measurement: PRESSURE METROLOGY</b>					
Measured quantity	Range	*Expanded Uncertainty ( ± )	Technique, Reference Standard, Equipment		
<b>Hydraulic / Lubricant / Liquid / Electronic Simulated Pressure</b> Source Mode	Active Lo Q-PRT (690 to 139.96) bar (1-2030) psi (0.009-14) MPa	1 – 200 psi	0.006 % of FS	<b>Reference Standards:</b> Electronic Dead Weight Tester (RPM4-EDWT, A140M), P-19 <b>Unit Under Test:</b> As mentioned below and electronic simulation-based pressure, transmitter / applications <b>Method Used:</b> PMD/TPP-38/ Fluke OEM DKD R-6-1 for CPB EURAMET PB/03 EAL-G26/EA-4/17	
		201 – 500 psi			
		501 – 1000 psi			
		1001 – 1500 psi			
		1501 – 2000 psi			
	Active Hi Q-PRT (137 to 1378) bar (2001-20000) psi (14-140) MPa	2001 – 3000 psi	0.009 % of FS		
		30001 – 5000 psi			
		5001 – 7000 psi			
		7001 – 10000 psi			
		10001 – 15000 psi			
		15001 – 20000 psi			
		20000 psi			
	<b>Hydraulic / Lubricant/Liquid Pressure</b> Source Mode	(1 to 690) bar	(10 – 500) psi		0.033 psi
			(501 – 1000) psi		0.034psi
or (10 to 10000) psi		(1001 – 5000) psi	0.052 psi		
		(5001 – 10000) psi	0.35 psi		
or (0.1 to 69) MPa		(690 to 1240) bar	0.15 % of FS		
		or (10001 to 18000) psi			
		or (69 to 124) MPa			
		(10001 – 11000) psi			
		(11001 – 13000) psi			
		(13001 – 15000) psi			
	(15001 – 18000) psi				

19-05-2025  
Date

Sd  
\_\_\_\_\_  
Director



## ACCREDITATION DOCUMENT

**F-06/02**  
**Issue Date: 18/08/2020**  
**Rev. No: 09**  
**LAB 001**

<b>Vacuum / Air / Gas / Low Pressure Source Mode</b>	(-1 to 20) bar or (-14 to 290) psi or (-100 to 2000) kPa	(-1 – 0.1) bar	0.001 bar	<p><b>Reference Standards:</b> Pneumatic Pressure Calibrator with DPI/RPG P-10 &amp; P-13 (Pneumatic Pressure Indicator), P-16, P-20 and P-21</p> <p><b>Unit Under Test:</b> Bourdon Tube / Diaphragm Pressure Gauge / Switch Digital Vacuum / Low Pressure Transducer / Gauges / Indicator / Switch / Adapter / Sensor / Valve / Snubber / Module Barometer / Manometer</p> <p><b>Method Used:</b> PMD/TPP-24 / Druck OEM DKD R-6-1 for CPB EURAMET PB/03 EAL-G26/EA-4/17</p>
		(0– 20) bar	0.041 bar	
<b>Barometric/ Atmospheric Pressure (Measure Mode)</b>	(0.70 – 1.10) bar	0.005 % of FS		<p><b>Reference Standards:</b> Barometer Fluke RPM4, BA100ks, P-12 and Druck DPI 142 P-11</p> <p><b>Unit Under Test:</b> Aneroid / Altimeter / Atmospheric meter Barometers and Manometers</p> <p><b>Method Used:</b> PMD/TPP-30 / Druck OEM / DPI K0343</p>
<b>Hydraulic / Lubricant/Liquid Gauge Pressure (Measure Mode)</b>	(1 to 690) bar or (10 to 10000) psi or (0.1 to 69) MPa	(0-500) psi	0.56 psi	<p><b>Reference Standards:</b> Hydraulic Dead Weight Tester / Pressure Balance P-01, P-01/R and P-09 DPI / PG P-15 and SPG P-01/G</p> <p><b>Unit Under Test:</b> Hydraulic / Lubricant DWT / PB, Pressure Balance / Pump / Source</p> <p><b>Method Used:</b> PMD/TPP-23 &amp; 29 / Druck OEM DKD R-6-1 for CPB</p>
		(501-1000) psi	0.56 psi	
		(1001-3000) psi	0.60 psi	
		(3001-5000) psi	0.90 psi	
		(5001-7000) psi	1.43 psi	
		(7001-10000) psi	1.98 psi	
<b>Pneumatic / Air / Gauge Pressure (Measure Mode)</b>	(-1 to 20) bar or (-14 to 290) psi or (-100 to 2000) kPa	-1 to 0 bar	0.001 bar	<p><b>Reference Standards:</b> Pneumatic Pressure Calibrator with DPI/RPG P-10 &amp; P-13, P-16, P-20 and P-21</p> <p><b>Unit Under Test:</b> Pneumatic / Air data calibrator / Vacuum / Low Pressure Pump</p> <p><b>Method Used:</b> PMD/TPP-24 / Druck OEM DKD R-6-1 for CPB</p>
		0 to 2 bar	0.002 bar	
		0 to 20 bar	0.041 bar	

### Field of measurement: THERMAL METROLOGY

Measured quantity	Range	*Expanded Uncertainty ( ± )	Technique, Reference Standard, Equipment
<b>Temperature (Sources)</b>	-80.00 °C	0.05 °C	<p><b>Reference Standards:</b> Black Stack Thermometer FLUKE, Working Standard PRT FLUKE,</p>
	0.00 °C	0.02 °C	
	100 °C	0.12 °C	

19-05-2025  
Date

\_\_\_\_\_  
Sd  
Director



## ACCREDITATION DOCUMENT

**F-06/02**  
**Issue Date: 18/08/2020**  
**Rev. No: 09**  
**LAB 001**

	200 °C	0.13 °C	R-Type Thermocouple 5649 FLUKE Zero Point Dry Well FLUKE Dry Block Calibrator (Model: Gemini700LRI) <b>Unit Under Test:</b> All types of Heat Sources <b>Method Used:</b> TMD/TPP-33
	300 °C	0.02 °C	
	400.0 °C	0.03 °C	
	500.0 °C	0.02 °C	
	600.0 °C	0.60 °C	
	700.0 °C	0.60 °C	
	800 °C	0.60 °C	
	1150 °C	2.20 °C	
Temperature (Measure)	-40 °C	0.26 °C	<b>Reference Standards:</b> Ultra Low Temperature Bath (Heart Scientific 7380) Liquid In Glass Thermometer (ASTM) Dry Block Calibrator ISOTECH GIMNI 700LRI Working Standard PRT (5628) Black Stack Digital Readout (1560) <b>Unit Under Test:</b> Liquid –In- Glass Thermometers <b>Method Used:</b> TMD/TPP-25
	0 °C	0.24 °C	
	10 °C	0.24 °C	
	30 °C	0.24 °C	
	50 °C	0.24 °C	
	80 °C	0.24 °C	
	100 °C	0.24 °C	
	150 °C	0.30 °C	
	200 °C	0.32 °C	
Temperature / Humidity (Sources)	20 %RH	0.6 %RH	<b>Reference Used:</b> RH Generator (Humi-Lab) NESLAB RTE Bath/Circulator (USA) Humidity and Temperature Probe OMEGAETTE (USA) Model: HH311 Fluke 971 <b>Unit Under Test:</b> All types of Thermo hygrometers <b>Method Used:</b> TMD/TPP-32
	30 %RH	0.6 %RH	
	40 %RH	0.6 %RH	
	50 %RH	0.6 %RH	
	60 %RH	0.6 %RH	
	70 %RH	0.7 %RH	
	80 %RH	0.8 %RH	
	18 °C	0.2 °C	
	20 °C	0.2 °C	
	22 °C	0.2 °C	
	24 °C	0.2 °C	
	26 °C	0.2 °C	
	28 °C	0.2 °C	
Temperature (Measure)	-80 °C	0.0025 °C	<b>Reference Used:</b> Working Standard PRT (5628) Black Stack Digital Readout (1560) Dry Block Calibrator ISOTECH GIMNI 700LRI Ultra-Low Temperature Bath (7380) <b>Unit Under Test:</b> All types of PRT/RTD <b>Method Used:</b> TMD/TPP-31
	0 °C	0.0010 °C	
	100 °C	0.0015 °C	
	200 °C	0.0015 °C	
	300 °C	0.0015 °C	
	400 °C	0.002 °C	
	500 °C	0.002 °C	
	600 °C	0.0035 °C	
Temperature (Sources)	0 °C	0.60 °C	<b>Reference Used:</b> Dry Block Calibrator (Model: Gemini700LRI) Ultra-Low Temperature Bath (Model:7380 USA), E-Type Thermocouple (Serial#:2H19 Japan) Black Stack Digital Readout (Serial#: A8B906, Fluke USA)
	200 °C	0.60 °C	
	400 °C	0.60 °C	

19-05-2025  
Date

Sd  
\_\_\_\_\_  
Director



## ACCREDITATION DOCUMENT

**F-06/02**  
**Issue Date: 18/08/2020**  
**Rev. No: 09**  
**LAB 001**

	600 °C	0.60 °C	Furnace (9112B, Fluke USA) Temperature/Pressure Calibrator (Model#: 525BFluke USA), Working Standard PRT (Model #: 5628) R-Type Thermocouple Serial#:4996), Muffle Furnace (Model#:KE-6HK1200-3) Zero Point Dry Well (Model#:9101) S-Type Thermocouple (Model#:5650) <b>Unit Under Test:</b> S-Type, R-Type and K-Type Thermocouple <b>Method Used:</b> TMD/TPP-26
	800 °C	0.60 °C	
	1000 °C	0.60 °C	
	1100 °C	2.00 °C	

### Field of measurement: TIME AND FREQUENCY METROLOGY

Measured quantity	Range	*Expanded Uncertainty ( ± )	Technique, Reference Standard, Equipment
<b>Frequency (Source)</b>	10 Hz to 500 Hz @0dBm	7.80 E-03	<b>Reference Standards:</b> a) Cesium Frequency Standard 5071A b) Precision Test Systems RFS10, Rb Frequency Standard c) Marconi Instrument, AM/FM signal Generator 2024 (Source) d) SMF100 Signal Generator <b>Unit Under Test:</b> Agilent Universal Frequency Counter/Timer 53230A (Measure) Microwave Frequency Counter 53152A <b>Method Used:</b> Direct Frequency Comparison (T&FMD/TPP-04)
	501 Hz to 10 kHz @0dBm	7.80 E-03 to 2.30 E-05	
	11 Hz to 500 kHz @0dBm	2.30 E-05	
	501 kHz to 1 MHz @0dBm	2.30 E-05 to 3.20 E-08	
	2 MHz to 100 MHz @0dBm	3.20 E-08 to 4.10 E-07	
	101 MHz to 500 MHz @0dBm	4.10 E-07 to 5.00 E-07	
	501 MHz to 1 GHz @0dBm	5.00 E-07 to 5.00 E-09	
	2 GHz to 10 GHz @0dBm	5.00 E-09 to 6.20 E-09	
	11 GHz to 20 GHz @0dBm	6.20 E-09 to 6.90 E-09	
	21 GHz to 30 GHz @0dBm	6.90 E-09 to 1.00 E-08	
	31 MHz to 40 GHz @0dBm	1.00 E-08 to 1.40 E-08	
	<b>Time base 10 MHz</b>	<b>3.4 E-08</b>	
<b>Frequency (Measure)</b>	10 Hz - 500 Hz @ 0 dBm	7.80 E-03	<b>Reference Standards:</b> Cesium Frequency Standard 5071A Agilent Universal Frequency Counter/Timer 53230A Microwave Frequency Counter 53152A <b>Unit Under Test:</b> Marconi Instrument, AM/FM signal Generator 2024 SMF100 Signal Generator Power Meter E4416A Agilent
	501 Hz - 10 kHz @ 0 dBm	7.80 E-03 to 2.40 E-05	
	11 kHz - 500 kHz @ 0 dBm	2.40 E-05	
	501 kHz - 1 MHz @ 0 dBm	2.40 E-05 to 3.20 E-08	
	2 MHz - 100 MHz @ 0 dBm	3.20 E-08	
	101 MHz - 500 MHz @ 0 dBm	3.20 E-08 to 3.50 E-08	

19-05-2025  
Date

\_\_\_\_\_  
Sd  
Director





## ACCREDITATION DOCUMENT

**F-06/02**  
**Issue Date: 18/08/2020**  
**Rev. No: 09**  
**LAB 001**

	501 MHz - 1 GHz @ 0 dBm	3.50 E-08 to 4.80 E-09	<b>Method Used:</b> Direct Frequency Comparison with Frequency Counter (T&FMD/TPP-01)
	2 GHz - 10 GHz @ 0 dBm	4.80 E-09 to 6.20 E-09	
	11 GHz - 20 GHz @ 0 dBm	6.20 E-09 to 6.90 E-09	
	21 GHz - 30 GHz @ 0 dBm	6.90 E-09 to 1.00 E-08	
	31 GHz - 40 GHz @ 0 dBm	1.00 E-08 to 1.40 E-08	
	<b>Time base 10 MHz</b>	<b>3.20 E-08</b>	
<b>Power Level (Measure)</b>	-5 dBm to +20 dBm @ 50 MHz	3.40 E-08	<b>Method Used:</b> Direct Frequency Comparison with Power Meter (T&FMD/TPP-08)
<b>Time Interval Measure</b>	1s to 60s (1min)	4.80E-01	<b>Reference Standards:</b> Precision Test Systems GPS10RBN <b>Unit Under Test:</b> Casio, HS-60W-IDF, Q&Q Digital Stopwatch <b>Method Used:</b> Direct Frequency Comparison (T&FMD/TPP-02)
	1s to 1800s (1/2 hr)		
	1s to 3600s (1hrs)		
	1s to 7200s (2 hrs)		

### Field of measurement: ELECTRICAL METROLOGY

#### A. (Measure Mode Scope)

Measured quantity	Range	*Expanded Uncertainty ( ± )	Technique, Reference Standard, Equipment
<b>DC Voltage</b>	0 - 100 mV	0.000059 mV	<b>Reference Standards:</b> 3458-Agilent, 34401 – Agilent, Reference Multimeter 8508 - A, Fluke, Digital multi-meters High voltage divider DVR 150 <b>Unit Under Test:</b> Fluke-5720-A, 9100-Wavetek, Multifunction Calibrator, High voltage source etc. <b>Method Used:</b> EMD/TPP-09
	100.00001 mV - 1 V	0.000070 V	
	1.000001 - 10 V	0.0000021 V	
	10.00001 - 100 V	0.0000072 V	
	100.0001 - 1000 V	0.000075 V	
	1000.001 V - 150 kV	0.00055 kV	
<b>DC Current</b>	0 - 100 μA	0.00022 μA	<b>Reference Standard:</b> 3458-Agilent, 34401 – Agilent, Reference Multimeter 8508 - A, Fluke, Digital Multi-meters, Clamp meter Radian TX-21 <b>Unit Under Test:</b> Fluke-5720-A, 9100-Wavetek Multifunction Calibrator (Source) <b>Method Used:</b> EMD/TPP-09
	100.0001 μA - 1 mA	0.00064 mA	
	1.000001 - 10 mA	0.000021 mA	
	10.00001 - 100 mA	0.000040 mA	
	100.01 mA - 3 A	0.000026 A	
	3.00001 - 20 A	0.00050 A	
	20.00001 - 2000 A	1.5 %	
<b>AC Voltage</b>	0 - 10 mV, at 50Hz to 300 Hz	0.00046 mV	<b>Reference Standard:</b> 3458-Agilent, 34401 – Agilent, Reference Multimeter 8508 - A, Fluke,
	10.0001 - 100 mV, at	0.00046 mV	

19-05-2025  
Date

\_\_\_\_\_  
Sd  
Director



## ACCREDITATION DOCUMENT

**F-06/02**  
**Issue Date: 18/08/2020**  
**Rev. No: 09**  
**LAB 001**

	50Hz to 300 Hz		Digital Multi-meters (Measure) High Voltage Divider <b>Unit Under Test:</b> Multifunction Calibrator Fluke-5720-A, High Voltage Source etc. <b>Method Used:</b> EMD/TPP-09
	100.00001 mV - 1 V, at 50 to 500 Hz	0.00016 V	
	1.00001 - 10 V, at 50 Hz to 10 kHz	0.00007 V	
	10.00001 - 100 V, at 50 Hz to 10 kHz	0.00040 V	
	100.0001 - 1000 V, at 50 Hz to 10 kHz	0.0030 V	
	1kV to 150kV, at 50Hz to 1kHz	0.016 kV	
<b>AC Current</b>	0 - 200 $\mu$ A, at 10 Hz to 1 kHz	0.0020 $\mu$ A	<b>Reference Standard:</b> 3458-Agilent, 34401 – Agilent, Reference Multimeter 8508 - A, Fluke, Digital Multi-meters (Measure), Clamp meter <b>Unit Under Test:</b> Fluke-5720-A, Multifunction Calibrator (Source), High Voltage Source etc. <b>Method Used:</b> EMD/TPP-09
	200.001 $\mu$ A - 10 mA, at 10 Hz to 1 kHz	0.0013 mA	
	10.0001 - 100 mA, at 10 Hz to 1 kHz	0.00017 mA	
	100.001 mA - 3 A, at 10 Hz to 1 kHz	0.00014 A	
	3.00001 - 20 A, at 10 Hz to 1 kHz	0.0019 A	
	20.00001 - 2000 A, at 50 Hz to 1 kHz	0.82 A	
<b>Resistance</b>	0 - 10 $\Omega$	0.000012 $\Omega$	<b>Reference Standard:</b> 3458-Agilent, 34401 – Agilent, Reference Multimeter 8508 - A, Fluke, Digital Multi-meters (Measure) <b>Unit Under Test:</b> Multifunction Calibrator Fluke-5720-A, 9100-Wavetek Multifunction Calibrator, Fixed Standard Resistors. AC/DC Shunts etc. (Source) <b>Method Used:</b> EMD/TPP-09, EMD/TPP-10, EMD/TPP-11, EMD/TPP-12
	10.00001 - 100 $\Omega$	0.000016 $\Omega$	
	100.0001 $\Omega$ - 1 k $\Omega$	0.0000065 k $\Omega$	
	1.0000001 - 100 k $\Omega$	0.0000065 k $\Omega$	
	100.00001 K $\Omega$ - 1 M $\Omega$	0.0000077 M $\Omega$	
	1.00001 - 100 M $\Omega$	0.0000087 M $\Omega$	
	100.001 M $\Omega$ - 2 G $\Omega$	0.000031 G $\Omega$	
	2.00001 - 20 G $\Omega$	0.000034G $\Omega$	
<b>Capacitance</b>	0.00 - 10 pF	0.058 pF	<b>Reference Standard:</b> PM-6306 Fluke, Reference Digital Capacitance Meter (Measure) <b>Unit Under Test:</b> Universal Calibration System (Sources), Discrete Standard Capacitors <b>Method Used:</b> EMD/TPP-13
	10.001 - 1000 pF	0.041 pF	
	1.000 – 1000 nF	0.67 nF	
	1.000 - 1000 $\mu$ F	0.00067 $\mu$ F	
<b>Inductance</b>	100 – 1000 $\mu$ H	0.067 $\mu$ H	<b>Reference Standard:</b> PM 6306 Fluke, Reference Digital Capacitance Meter (Measure) Digital Inductance / LCR Meter PM-6304 Fluke, (Measure) <b>Unit Under Test:</b> Universal Calibration System (Sources), Discrete Standard Inductors <b>Method Used:</b> EMD/TPP-14
	1 - 10 mH	0.00010 mH	
	10.001 - 90 mH	0.0076 mH	
	90.01 - 1000 mH	0.0095 mH	
	1 – 10 H	0.00030 H	

19-05-2025  
Date

Sd  
Director



**ACCREDITATION  
DOCUMENT**

**F-06/02**  
**Issue Date: 18/08/2020**  
**Rev. No: 09**  
**LAB 001**

<b>AC/DC Power</b>	50 mW – 87 kW (Direct) Up to 96 MW (with Sensors)	0.13 mW to 0.0067 kW 9.0 mW (with Sensors)	<p><b>Reference Standard:</b>  Three Phase Power Network Tester/ Analyzer (TE30)  Three Phase 4 Channel Power Analyzer (PA2203A)  Comparator COM5003, Zeera, Germany</p> <p><b>Unit Under Test:</b>  Electrical Power Standard (6105A),  Three Phase Current &amp; Voltage Source (MT551)  Reference Standard Universal Calibration System (Source)</p> <p><b>Method Used:</b>  EMD/TPP-16</p>
--------------------	--	---	--

**Field of measurement: ELECTRICAL METROLOGY**  
**B. (Source Generation Scope)**

Measured quantity	Range	*Expanded Uncertainty (±)	Technique, Reference Standard, Equipment
<b>DC Voltage</b>	1.018 V	0.20 μV	<p><b>Reference Standard:</b>  Fluke – 732-B, DC Voltage Standard (Source)</p> <p><b>Unit Under Test:</b>  3458-Agilent, Reference Multimeter 8508 - A, Fluke, Digital Multi-meter MFC 5720 A, 5700 A (Measure)</p> <p><b>Method Used:</b>  EMD/TPP-10</p>
	10 V	0.40 μV	
	220 mV	0.00020 mV	<p><b>Reference Standard:</b>  Fluke – 5720-A Multifunction Calibrator (Source)</p> <p><b>Unit Under Test:</b>  All Types of Digital / Analog Multi-meter 3458-AGILENT, 8508-A, Fluke, 45-Fluke, etc.</p> <p><b>Method Used:</b>  EMD/TPP-10</p>
	2.2 V	0.000001 V	
	11 V	0.000011 V	
	22 V	0.000020 V	
	220 V	0.00020 V	
1100 V	0.0020 V		
<b>DC Current</b>	220 μA	0.15 μA	<p><b>Reference Standard:</b>  Fluke – 5720-A Multifunction Calibrator (Source)</p> <p><b>Unit Under Test:</b>  All Types of Digital / Analog Multi-meters 3458-Agilent, 8508- A, Fluke, 45-Fluke, Clamp meter, etc. (Measure)</p> <p><b>Method Used:</b>  EMD/TPP-10</p>
	1.2 mA	0.0000060 mA	
	22 mA	0.000060 mA	
	220 mA	0.00060 mA	
	2.2 A	0.000013 A	
<b>AC Voltage</b>	20 A, 1000 A (via current coil)	0.0041 A 0.0044 A	<p><b>Reference Standard:</b>  Fluke – 5720- A Multifunction Calibrator etc. (Source)</p> <p><b>Unit Under Test:</b></p>
	2.2 mV, at 10 Hz to 100 kHz	0.000042 mV	
	22 mV, at 10 Hz to 100 kHz	0.00040 mV	

19-05-2025  
Date

Sd  
Director



## ACCREDITATION DOCUMENT

**F-06/02**  
**Issue Date: 18/08/2020**  
**Rev. No: 09**  
**LAB 001**

	220 mV, at 10 Hz to 100 kHz	0.00034 mV	<p>All Types of Digital / Analog Multi- meters i.e. 3458 -Agilent, 8508 - A, Fluke, 45-Fluke etc. (Measure)</p> <p><b>Method Used:</b> EMD/TPP-10 EMD/TPP-08</p>
	2.2 V, at 10 Hz to 100 kHz	0.0000020 V	
	22 V, at 10 Hz to 100 kHz	0.000025 V	
	220 V, at 10 Hz to 100 kHz	0.00022 V	
	750 V, at 40 Hz to 10 kHz	0.00022 V	
	1100 V, at 40 Hz to 10 kHz	0.0025 V	
<b>AC Current</b>	220 $\mu$ A, at 10 Hz to 10 kHz	0.00033 $\mu$ A	<p><b>Reference Standard:</b> Fluke – 5720-A Multifunction Calibrator - 9100 (Source)</p> <p><b>Unit Under Test:</b> All Types of Digital / Analog Multi-meter i.e. 3458-Agilent, 8508-A, Fluke, 45-Fluke, clamp meters etc. (Measure)</p> <p><b>Method Used:</b> EMD/TPP-10</p>
	2.2 mA, at 10 Hz to 10 kHz	0.0000022 mA	
	22 mA, at 10 Hz to 10 kHz	0.000025 mA	
	220 mA, at 10 Hz to 10 kHz	0.00024 mA	
	2.2 A, at 20 Hz to 10 kHz	0.0000037 A	
	20 A, at 40 Hz to 10 kHz 1000 A, at 40 Hz to 1 kHz (via current coil)	0.000015 A	
<b>Resistance</b>	1 $\Omega$	0.00000073 $\Omega$	<p><b>Reference Standard:</b> Fluke – 5720-A Multifunction Calibrator, High Resistance Decade Box RH9A-5</p> <p><b>Unit Under Test:</b> All Types of Digital / Analog Multi-meter i.e. 3458-Agilent, 8508-A, Fluke, 45-Fluke, etc. (Measure)</p> <p><b>Method Used:</b> EMD/TPP-10, EMD/TPP-11, EMD/TPP-12</p>
	10 $\Omega$	0.000267 $\Omega$	
	100 $\Omega$	0.000089 $\Omega$	
	1 k $\Omega$	0.0000018 k $\Omega$	
	10 k $\Omega$	0.0000050 k $\Omega$	
	100 k $\Omega$	0.0010 k $\Omega$	
	1 M $\Omega$	0.000010 M $\Omega$	
	10 M $\Omega$	0.00020 M $\Omega$	
	100 M $\Omega$	0.028 M $\Omega$	
1 G $\Omega$	0.000022 G $\Omega$		
<b>Resistance</b>	0.0001 $\Omega$	60.00 $\mu\Omega$	<p><b>Reference Standard:</b> Hi-Accuracy Working Standard, 4-Terminal Standard Resistors 20-E/D to 28 E/D, H&amp;B Germany (Source)</p> <p><b>Unit Under Test:</b> 34420-Agilent Digital <math>\mu\Omega</math> Meter (Measure)</p> <p><b>Method Used:</b> EMD/TPP-11, EMD/TPP-12</p>
	10 k $\Omega$	0.008 $\Omega$	
<b>Capacitance</b>	0.1 $\mu$ F	0.000080 $\mu$ F	<p><b>Reference Standard:</b> Fixed Value Working Standard Capacitors (Source)</p> <p><b>Unit Under Test:</b> Digital Capacitance/ LCR Meter PM-6304 Fluke, (Measure)</p> <p><b>Method Used:</b> EMD/TPP-13</p>
	1 $\mu$ F	0.00010 $\mu$ F	
<b>AC Power</b>	0.02 W – 72 kW (Direct), Up to 1MW (with coil)	0.01 W to 0.16 kW	<p><b>Reference Standard:</b> 6105A Electrical Power Standard., MT551 Three Phase Voltage and Current Source</p>

19-05-2025  
Date

Sd  
\_\_\_\_\_  
Director



## ACCREDITATION DOCUMENT

**F-06/02**  
**Issue Date: 18/08/2020**  
**Rev. No: 09**  
**LAB 001**

			<b>Unit Under Test:</b> Power Meters / Analyzers etc. <b>Method Used:</b> EMD/TPP-17
<b>Inductance</b>	100 $\mu$ H	0.020 $\mu$ H	<b>Reference Standard:</b> Fixed Value Working Standard Inductors (Source)
	1 mH	0.060 mH	
	10 mH	0.050 mH	<b>Unit Under Test:</b> Digital Inductance/ LCR Meter PM-6304 Fluke (Measure)
	100 mH	0.010 mH	<b>Method Used:</b> EMD/TPP-14
	1 H	0.00040 H	

### Field of measurement: CHEMICAL METROLOGY

Measured quantity	Range	*Expanded Uncertainty ( $\pm$ )	Technique, Reference Standard, Equipment
<b>Electrical Conductivity (EC)</b>	1413 $\mu$ mhos / cm	5.2 $\mu$ mhos /cm	<b>Calibration of Conductivity meter</b> <b>Reference Standard:</b> NIST Traceable Conductivity standard solution <b>Unit Under Test:</b> Conductivity Meter <b>Method Used:</b> APHA, AWWA, WEF, USA, 2510B
	12.9 mmhos / cm	0.10 mmhos /cm	
<b>pH</b>	0 -14	0.02	<b>Reference Standard:</b> NIST Traceable pH buffers <b>Unit Under Test:</b> pH Meter <b>Method Used:</b> APHA, AWWA, WEF, USA, 2510B Standard Buffer solutions / Two - point calibration method

### Calibration Laboratory.

Onsite Accreditation Scope of **NATIONAL METROLOGY INSTITUTE OF PAKISTAN**  
 Plot No. 16, Sector H-9/1, Islamabad, Pakistan

Mobile laboratory (Onsite Calibration)

### Field of measurement: MASS METROLOGY

Measured quantity	Range	*Expanded Uncertainty ( $\pm$ )	Technique, Reference Standard, Equipment
<b>Mass</b>	1mg to 500 g	0.010 to 0.10 mg (depends upon readability/resolution of unit under-test)	<b>Calibration of Balance / Weighing machine</b> <b>Reference Standards:</b> E2 class standard masses (1 mg to 5000 g) <b>Method Used:</b> MMD/TPP-21/04
	500 g to 5000 g	0.10 to 1.0 mg (depends upon readability/resolution of unit under-test)	

19-05-2025  
Date

\_\_\_\_\_  
Sd  
Director



**ACCREDITATION  
DOCUMENT**

**F-06/02**  
**Issue Date: 18/08/2020**  
**Rev. No: 09**  
**LAB 001**

**Field of measurement: THERMAL METROLOGY**

Measured quantity	Range	*Expanded Uncertainty (±)	Technique, Reference Standard, Equipment
Temperature	Ambient to 1100 °C	0.6 °C to 2.0 °C	<b>Calibration of Thermocouples</b> Dry block Temperature Calibrator, Digital Readout, Reference Thermometers, Zero-point Dry Well <b>Method Used:</b> TMD/TPP-26
	Ambient to 650 °C	0.0025 °C to 0.0035 °C	<b>Calibration of PRT/SPRT</b> Reference Thermometers, Digital Readout, Zero-point Dry Well, Dry block Temperature Calibrator <b>Method Used:</b> TMD/TPP-31
	-80 °C to 1400 °C	0.01 °C to 2.20 °C	<b>Calibration of Low Temperature Bath/ Heat Sources (Oven, Incubator, Dry Block, Furnace)</b> PRT, S & R Type Thermometer <b>Method Used:</b> TMD/TPP-34
Temperature and Humidity	20 %RH to 80 %RH 18 °C to 30 °C	0.6 % RH to 0.8 %RH 0.2 °C	<b>Calibration of Thermo-hygrometers</b> RH Generator Humi-Lab. Humidity and Temperature Probe (OMEGAETTE, USA; Model: HH311) <b>Method Used:</b> TMD/TPP-32

**Field of measurement: ELECTRICAL METROLOGY**

Measured quantity	Range	*Expanded Uncertainty (±)	Technique, Reference Standard, Equipment
DC Voltage	100 mV	0.00085 mV	<b>Reference Standard:</b> Digital multi-meters High voltage divider DVR 150 <b>Unit under test:</b> Multifunction Calibrator, High voltage source
	100.00001 mV - 1 V	0.0000042 V	
	1.000001 - 10 V	0.000037 V	
	10.00001 - 100 V	0.00056 V	
	100.0001 - 1000 V	0.0056 V	
	1 kV – 150 kV	0.5%	
AC Voltage	100 mV – 1000 V at 50 Hz to 10 kHz	0.0039 mV - 0.028 V	
	1kV - 150kV at 50Hz to 1kHz	0.5%	
DC Current	1 mA – 20 A	0.00015 mA – 0.00025A	<b>Reference Standard:</b> 3458-Agilent, 34401- Agilent, 8508- A, Fluke, Digital Multi-meters, Clamp meter Radian TX-21 <b>Unit under test:</b> Multifunction Calibrator High Current Source
	20 A – 2000 A	1.5%	
AC Current	1 mA to 20 A, at 10 Hz to 1 kHz	0.0013 mA	
	20 A – 2000 A, at 50 Hz to 10 kHz	1.5%	

19-05-2025  
Date

Sd  
\_\_\_\_\_  
Director

	<b>ACCREDITATION DOCUMENT</b>	<b>F-06/02</b> <b>Issue Date: 18/08/2020</b> <b>Rev. No: 09</b> <b>LAB 001</b>
---	-----------------------------------	---

<b>Resistance</b>	1 $\Omega$ - 2 G $\Omega$	0.00017 $\Omega$ - 0.00026 G $\Omega$	<b>Reference Standard:</b> Digital multi-meters <b>Unit Under Test:</b> Resistance Calibrator Fixed / Decade Resistance Box
	2 G $\Omega$ - 20 G $\Omega$	0.0026 G $\Omega$	
<b>Capacitance</b>	1 nF – 1000 $\mu$ F	0.001 nF – 0.058 $\mu$ F	<b>Reference Standard:</b> PM-6306 Fluke <b>Unit Under Test:</b> Decade / Fixed Capacitance Standards
<b>Inductance</b>	100 $\mu$ H – 1 H	0.05 $\mu$ H – 0.0061 H	<b>Reference Standard:</b> PM-6306 Fluke, <b>Unit under test:</b> Digital / Fixed / Decade Inductance

\* **Expanded Uncertainty:**

- Expanded Uncertainty is the measurement uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . This measurement uncertainty is a value for which the laboratory has been accredited using the procedure that was the subject of assessment. In certificates issued under its accreditation scope an accredited laboratory is not permitted to quote an uncertainty that is smaller than the published uncertainty for respective ranges as given above.

19-05-2025  
Date

\_\_\_\_\_  
Sd  
Director