

	ACCREDITATION DOCUMENT	F-06/02 Issue Date: 18/08/2020 Rev. No: 09 LAB 036
---	-----------------------------------	---

Accreditation No: LAB 036

Awarded to

**Applied Physics Computers & Instrumentation Centre (APCIC),
Pakistan Council of Scientific & Industrial Research (PCSIR) Labs.
Complex. Lahore 54600, Pakistan.**

The scope of accreditation is in accordance with the standard specifications outlined in the following page(s) of this document. The accredited scope shall be visible and legible in areas such as customer service, sample-receiving section etc and shall not mislead its users.

The accreditation was first time granted on **24-08-2006** by Pakistan National Accreditation Council.

The laboratory complies with the requirements of **ISO/IEC 17025:2017**.

The accreditation requires regular surveillance, and is valid until **16-09-2025**.

The decision of accreditation made by Pakistan National Accreditation Council implies that the organization has been found to fulfill the requirements for accreditation within the scope.

The organization however, itself is responsible for the results of performed measurements/tests.

PAKISTAN NATIONAL ACCREDITATION COUNCIL

16-06-2025
Date

SD
Director General

	ACCREDITATION DOCUMENT	F-06/02 Issue Date: 18/08/2020 Rev. No: 09 LAB 036
---	-----------------------------------	---

Calibration Laboratory.

Accreditation scope of Applied Physics Computers & Instrumentation Centre (APCIC), Pakistan Council of Scientific & Industrial Research (PCSIR) Laboratories Complex, Lahore 54600, Pakistan.

Permanent laboratory premises ☒

Field of Measurement: Volume Measurement			
Measured Quantity	Range	*Expanded Uncertainty (+)	Technique, Reference Standard, Equipment
Glassware	1 mL	1.0 µL	Reference Used: F2 Class Mass Weighing Scale UUC: Pipettes, Burettes, Dispensers, Cylinders, Flasks, Beakers, Jugs etc. Method Used: ASTM E-542 LLC/APCIC/VCP/01 LLC/APCIC/VCP/02
	2 mL	1.6 µL	
	5 mL	5.0 µL	
	10 mL	8.0 µL	
	25 mL	2.0 µL	
	25.1 mL to 100.0 mL	7.8 µL	
	100.1 mL to 500.0 mL	1.1 mL	
	500.0 to 1000.0 mL	6.0 mL	
	1.0 L to 5.0 L	5.0 mL	
Micropipette	20 µL to 1000 µL	0.25 µL	
Field of Measurement: Masses and Weighing Balances			
Weighing Scales	1.0 mg to 220.0000 g 0.1 g to 20.0000 kg 0.10 kg to 260.0 kg	0.10 mg 0.10 g 0.10 kg	Reference Used: E2 Class Mass F2 Class Mass UUC: Class I & Below F1 Class & Below
Masses	1.0 mg to 500.0 mg 1.0000 g to 200.0000 g 0.5000 kg to 20.0000 kg 30.000 kg to 50.000 kg	0.050 mg 0.060 mg 0.080 g 4.6 g	Method Used: OIML , R111-1 R111-2, NIMT CP-301 LLC/APCIC/MCP/01 LLC/APCIC/MCP/02
Field of Measurement: Temperature Measurement			
Temperature	- 50.00 °C to 50.00 °C	0.050 °C	Reference Used: Reference Thermometer with RTD Probe ,
	50.10 °C to 400.0 °C	0.20 °C	

16-06-2025

Date

SD

Director

	ACCREDITATION DOCUMENT	F-06/02 Issue Date: 18/08/2020 Rev. No: 09 LAB 036
---	-----------------------------------	---

	400.1 °C to 800.0 °C	0.30 °C	Precision k-type Thermocouple , Temperature Controller, Dry Well Calibrator UUC: Digital Thermometer Liquid in glass Thermometer Method Used: LLC/APCIC/TCP/01 LLC/APCIC/TCP/02
Temperature	-50.0 °C to 100.0 °C 100.1 °C to 350 °C	0.20 °C 0.28 °C	

Field of Measurement: Temperature Source

Temperature	- 50.0 °C to 100.0 °C 100.1 °C to 350.0 °C 350.1 °C to 600 °C	0.18 °C 0.20 °C 0.24 °C	Reference Used: Reference Thermometer with RTD Probe Precision Thermometer with k-type UUC: Thermocouple, Temperature Controllers Dry Block Calibrator / Temperature Calibrator Environmental Chambers / Oven Muffle Furnace Method Used: LLC/APCIC/TCP/01
Temperature	- 40.0 °C to 100.0 °C 100.1 °C to 300.0 °C	0.18 °C 0.22 °C	
Temperature	200 °C to 1000 °C	0.65 °C	

Temperature Measurement by Simulation Method

Temperature	- 100 °C to 800 °C	0.20 °C	Reference Used: Portable Calibrator, Fluke 8508A Reference Multimeter UUC: RTD Pt 100, Thermocouple Type k and J Method Used: LLC/APCIC/TCP/01
Temperature	- 200 °C to 1200 °C	0.24 °C	
Temperature	- 200 °C to 1200 °C	0.24 °C	

Field of Measurement: Temperature & Humidity Measurement

Source	10.0 °C to 40.0 °C 30 %RH to 80 %RH	0.45 °C 2.9 %RH	Reference Used: Thermo-hygrometer UUC: Humidity Chamber, Method used: LLC/APCIC/TCP/03
Measurement	10 °C to 40 °C 30 %RH to 80 %RH	0.45 °C 2.9 %RH	

16-06-2025

Date

SD

Director

	ACCREDITATION DOCUMENT	F-06/02 Issue Date: 18/08/2020 Rev. No: 09 LAB 036
---	-----------------------------------	---

Field of Measurement: Pressure Measurement			
Pressure	0.01 to 10.00 psi	0.31 psi	Reference Used: Pressure Calibrator Dead Weight Tester & Pressure Guages UUC:
	10.01 to 50.00 psi	0.36 psi	
	50.01 to 100.00 psi	0.37 psi	
	100.01 to 250.00 psi	0.38 psi	
Pressure	50 psi to 500 psi	1.4 psi	Pressure Gauge (Pneumatic , Vacuum & Hydraulic) Pressure calibrator , Dead weight tester Method Used: DKD-R 6-1 LLC/APCIC/PCP/01 LLC/APCIC/PCP/02
	550 psi to 1000 psi	1.9 psi	
	2000 psi to 5000 psi	5.7 psi	
	5000 psi to 8000 psi	8.8 psi	
	100 mm of Hg to 200 mm of Hg	12.31 mm of Hg	
	220 mm of Hg to 500 mm of Hg	12.31 mm of Hg	
	520 mm of Hg to 600 mm of Hg	12.31 mm of Hg	
Field of Measurement: Dimensional Measure			
General Dimension measurements Length, Diameter, Thickness and Depth of Industrial Artifacts	0.001 mm to 25.000 mm	0.40 μm	Reference Used: Gauge Block Set, Micrometer, Vernier Caliper, Line Length Standard, Measuring Tape
	25.10 mm to 100.00 mm	1.0 μm	
	100.01 mm to 300.00 mm	1.0 μm	
	1.0 cm to 100.0 cm	0.10 cm	Method Used: EAL-G 21 SOP # 10 & 12 (NIST) LLC/APCIC/DCP/01 LLC/APCIC/DCP/02 LLC/APCIC/DCP/03 LLC/APCIC/DCP/04 LLC/APCIC/DCP/05
	100.1 cm to 500.0 cm	0.10 cm	
Field of Measurement: Frequency			
Frequency Generation	10.0 Hz to 100.0 Hz	0.010 Hz	Reference Used : Universal Frequency Counter UUC:
	1.000 kHz to 100.00 kHz	0.039 Hz	
	1.00 MHz to 100.0 MHz	0.060 kHz	
Frequency Measurement	10.0 Hz to 100.0 Hz	0.010 Hz	Function Generator Frequency Counter, Digital Oscilloscope Method Used: LLC/APCIC/FCP/01
	1.000 KHz to 100.00 KHz	0.080 Hz	
	1.00 MHz to 100.00 MHz	0.060 kHz	
Field of Measurement: RPM Measurement			

16-06-2025
Date

SD
Director

	ACCREDITATION DOCUMENT	F-06/02 Issue Date: 18/08/2020 Rev. No: 09 LAB 036
---	-----------------------------------	---

Tachometers / RPM Measurement	30.0 RPM to 300.0 RPM 300.1 RPM to 6000.0 RPM 6000.1 RPM to 15000.0 RPM 15000.1 RPM to 30000.0 RPM	0.50 RPM 0.72 RPM 1.0 RPM 2.5 RPM	Reference Used: Universal Frequency Counter UUC: Function Generator, Digital Tachometers Method Used: LLC/APCIC/RCP/01
-------------------------------	---	--	--

Field of Measurement: Time Interval Measurement			
Stop Watch	10 s to 3600 s 3601 s to 7200 s	0.49 s 0.75 s	Reference Used: Universal Frequency Counter UUC: Function Generator, Frequency Counter, Stop Watches, Method Used: LLC/APCIC/WCP/01
Timer	10 s to 7200 s	0.75 s	

Field of Measurement: Electrical Parameters			
DC Voltage	1.00 mV to 100.00 mV 1.000 V to 10.000 V 10.001 V to 100.00 V 100.01 V to 1000.0 V	0.61 μ V 0.011 mV 0.11 mV 56 mV	Reference Used: Reference Multimeter Fluke 8508A UUC: Inmel Calibrator Clamp Meter Standard Resistors Method Used: LLC/APCIC/ECP/01 LLC/APCIC/ECP/02
AC Voltage @ 50 Hz	100.0 mV to 1.000 V 1.001 V to 10.000 V 10.001 V to 100.00 V 100.01 V to 500.0 V 500.1 V to 1000 V	0.46 mV 0.11 mV 2.6 mV 11 mV 51 mV	
AC Current @ 50 Hz	1.00 mA to 10.00 mA 10.01 mA to 100.0 mA 1.000 A to 10.000 A	0.76 μ A 3.6 μ A 4.8 mA	
AC Current (Clamp on) @ 50 Hz	5.0 A to 10.0 A 10.1 A to 100.0 A 100.1 A to 500.0 A 500.1 A to 800 A	0.20 A 0.52 A 1.1 A 2.0% A	
DC Current	1.00 mA to 10.00 mA 10.01 mA to 100.0 mA 1.000 A to 10.000 A	2.2 μ A 5.6 μ A 4.4 mA	
DC Current (Clamp on)	5.0 A to 10.0 A 10.1 A to 100.0 A 100.1 A to 500.0 A 500.1 A to 800 A	0.30 A 0.33 A 1.3 A 2.0% A	
Resistance	1.00 Ω to 10.00 Ω	0.013 Ω	

16-06-2025

Date

SD

Director

	ACCREDITATION DOCUMENT	F-06/02 Issue Date: 18/08/2020 Rev. No: 09 LAB 036
---	-----------------------------------	---

	10.01 Ω to 100.00 Ω	0.015 Ω	
	100.00 Ω to 1.000 k Ω	0.015 Ω	
	1.001 k Ω to 10.00 k Ω	0.10 Ω	
	10.01 k Ω to 100.00 K Ω	4.7 Ω	
	100.01 k Ω to 1.000 M Ω	0.54 K Ω	
	1.001 M Ω to 10.00 M Ω	8.7 k Ω	
	10.01 M Ω to 100.0 M Ω	15 k Ω	
Insulation Resistance @ 250 V to 1000 V	100.1 M Ω to 1.000 G Ω	5.0 M Ω	
Low Resistance Measurement	10.0 m Ω to 100.0 m Ω	0.14 m Ω	

Field of Measurement: AC Power (Single Phase)

AC Power @ 50 Hz	10.00 W to 100.0 W 100.1 W to 500.0 W 500.1 W to 1000.0 W 1000.1 W to 5000.0 W	0.12 W 0.24 W 0.42 W 1.0 W	Reference Used: Multimeter Fluke 8508A UUC: Power Meter Clamp Meter Inmel 33 Calibrator Method Used: LLC/APCIC/ECP/01 LLC/APCIC/ECP/02
---------------------	---	-------------------------------------	--

Field of Measurement: Spectrophotometer

Wavelength Accuracy	525.5 nm	0.78 nm	Reference Used: SS-1 Spectronics Standard Filters Thermo Spectronics USA UUC: Spectrophotometers Method Used: Thermo Scientific, USA LLC/APCIC/SCP/01
Transmittance @ 590nm & 412 nm	6.13%T to 10.4 %T	0.011 %T	
Absorbance @ 590nm & 412 nm	0.990 A to 1.209 A	0.043 A	

Field of Measurement: pH Measurement

pH Meter	4.00 pH to 10.00 pH @ 25 °C	0.010 pH	Reference Used: HANNA pH Buffers, pH Meter UUC: pH Meters
----------	--------------------------------	----------	--

16-06-2025

Date

SD

Director

	ACCREDITATION DOCUMENT	F-06/02 Issue Date: 18/08/2020 Rev. No: 09 LAB 036
---	-----------------------------------	---

		<u>Method Used:</u> ASTM D 1293-12 LLC/APCIC/SCP/02
--	--	--

*** Expanded Uncertainty:**

- Expanded Uncertainty is the measurement uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. This measurement uncertainty is a value for which the laboratory has been accredited using the procedure that was the subject of assessment. In certificates issued under its accreditation scope an accredited laboratory is not permitted to quote an uncertainty that is smaller than the published uncertainty for respective ranges as given above.

16-06-2025
Date

SD
Director