

	<b>ACCREDITATION DOCUMENT</b>	<b>F-06/02 Issue Date: 10/08/15 Rev. No: 07 LAB 057</b>
---	-----------------------------------	---

**Accreditation No: LAB 057**

**Awarded to**

**(Analytical Laboratory)  
Intertek Pakistan (Pvt.) Limited  
Plot No. 1-5/11-A, Sector-5, Korangi Industrial Area,  
Karachi 74900, PAKISTAN**

The scope of accreditation is in accordance with the standard specifications outlined in the following page(s) of this document. The accredited scope shall be visible and legible in areas such as customer service, sample-receiving section etc and shall not mislead its users.

The accreditation was first time granted on **24-08-2012** by Pakistan National Accreditation Council.

The laboratory complies with the requirements of **ISO/IEC 17025:2005**.

The accreditation requires regular surveillance, and is valid until **05-10-2021**.

The decision of accreditation made by Pakistan National Accreditation Council implies that the organization has been found to fulfill the requirements for accreditation within the scope.

The organization however, itself is responsible for the results of performed measurements/tests.

**PAKISTAN NATIONAL ACCREDITATION COUNCIL**

**30-04-2019**

Date

\_\_\_\_\_  
Director General

	<b>ACCREDITATION DOCUMENT</b>	<b>F-06/02</b> <b>Issue Date: 10/08/15</b> <b>Rev. No: 07</b> <b>LAB 057</b>
---	-----------------------------------	---

**Testing Laboratory.**

Accreditation Scope of Intertek Pakistan (Pvt.) Limited. Plot No. 1-5/11-A, Sector-5, Korangi Industrial Area, Karachi 74900, PAKISTAN.

Permanent laboratory premises

Materials/Products tested	Testing field (e.g. environmental testing or mechanical testing)	Types of test/ Properties measured	Reference to standardized method (e.g. ISO 14577-1:2003)/ Internal method reference
Lube Oil	Petrochemicals	Acid Number	ASTM D 974
		Acid Number	ASTM D 664
		Base Number	ASTM D 2896
		Viscosity @ 40 <sup>0</sup> C	ASTM D 445
		Viscosity @ 100 <sup>0</sup> C	ASTM D 445
		Color	ASTM D 1500
		Ca Content	ASTM D 4628
		Zn Content	ASTM D 4628
		Density @ 15 <sup>0</sup> C	ASTM D 4052
		Density @ 20 <sup>0</sup> C	ASTM D 4052

**30-04-2019**

Date

\_\_\_\_\_  
Director