



**ACCREDITATION
DOCUMENT**

**F-06/02
Issue Date: 18/08/2020
Rev. No: 09
LAB 103**

Accreditation No: LAB 103

Awarded to

**ADVANCE ENGINEERING & RESEARCH ORGANIZATION
(AERO) CALIBRATION LAB.
LUB THATTO, HASSAN ABDAL, PAKISTAN.**

The scope of accreditation is in accordance with the standard specifications outlined in the following page(s) of this document. The accredited scope shall be visible and legible in areas such as customer service, sample-receiving section etc and shall not mislead its users.

The accreditation was first time granted on **15-03-2016** by Pakistan National Accreditation Council.

The laboratory complies with the requirements of **ISO/IEC 17025:2017**.

The accreditation requires regular surveillance, and is valid until **14-03-2025**.

The decision of accreditation made by Pakistan National Accreditation Council implies that the organization has been found to fulfill the requirements for accreditation within the scope.

The organization however, itself is responsible for the results of performed measurements/tests.

PAKISTAN NATIONAL ACCREDITATION COUNCIL

14-03-2024
Date

SD
Director General



**ACCREDITATION
DOCUMENT**

F-06/02
Issue Date: 18/08/2020
Rev. No: 09
LAB 103

Calibration Laboratory.

Advance Engineering & Research Organization (AERO) Calibration Lab.

Permanent laboratory premises **X**

Field of Measurement:				
Measured Quantity	Range		*Expanded Uncertainty (+)	Technique, Reference Standard, Equipment
DC Voltage (Source Mode)	0 to 329.9999 mV		4.3E-04 mV to 2.4E-03 mV	Fluke Warranted METCAL Procedures, Fluke 5522A Multi Product Calibrator
	330 mV to 3.299999 V		1.4E-03 mV to 1.4E-05 V	
	3.3 V to 32.99999 V		6.0E-04 V to 6.1E-04 V	
	33 V to 329.9999 V		4.5E-04 V to 4.1E-03 V	
	330 V to 1020 V		1.8E-03 V to 4.2E-03 V	
DC Current (Source Mode)	0 to 329.999 μ A		8.3E-04 μ A to 2.3E-02 μ A	
	330 μ A to 3.29999 mA		1.4E-02 μ A to 1.2E-04 mA	
	3.3 mA to 32.9999 mA		1.1E-04 mA to 6.4E-04 mA	
	33 mA to 329.999 mA		2.5E-03 mA to 2.5E-02 mA	
	330 mA to 1.09999 A		5.5E-02 mA to 2.3E-04 A	
	1.1 A to 2.99999 A		2.3E-04 A to 6.0E-04 A	
	3 A to 10.9999 A		3.5E-04 A to 6.2E-04 A	
	11 A to 20.5 A		6.2E-04 A to 8.2E-04 A	
AC Current (Source Mode)	29.00 μ A to 329.99 μ A	10 Hz to 30 kHz	8.2E-04 μ A to 1.7E-01 μ A	

14-03-2024
Date

Sd
Director



ACCREDITATION DOCUMENT

F-06/02
Issue Date: 18/08/2020
Rev. No: 09
LAB 103

	0.33 mA to 3.29999 mA		2.2E-04 mA to 2.0E-02 mA	
AC Current (Source Mode)	3.3 mA to 32.9999 mA	10 Hz to 30 kHz	2.0E-03 mA to 1.0E-01 mA	Fluke Warranted METCAL Procedures, Fluke 5522A Multi Product Calibrator
	33 mA to 329.9999 mA		1.3E-03 mA to 3.1E-01 mA	
	0.33 A to 1.09999 A		1.0E-04 A to 6.6E-03 A	
	1.1 A to 2.99999 A		1.7E-03 A to 5.0E-03 A	
	3 A to 10.9999 A		1.4E-03 A to 5.6E-03 A	
	11 A to 20.5 A		3.1E-03 A to 5.6E-03 A	
AC Voltage (Source Mode)	1.0 mV to 32.999 mV	10 Hz to 500 kHz	3.4E-03 mV to 1.8E-02 mV	Fluke Warranted METCAL Procedures, Fluke 5522A Multi Product Calibrator
	33 mV to 329.999 mV		7.9E-03 mV to 7.9E-02 mV	
	0.33 V to 3.29999 V		7.7E-05 V to 7.9E-04 V	
	3.3 V to 32.9999 V		8.8E-03 V to 8.8E-02 V	
	33 V to 329.999 V		8.9E-03 V to 8.8E-02 V	
	330 V to 1020 V		3.0E-02 V to 1.4E-01 V	
Resistance (Source Mode)	0 Ω to 10.9999 Ω		1.4E-03 Ω to 1.5E-04 Ω	Fluke Warranted METCAL Procedures, Fluke 5522A Multi Product Calibrator
	11 Ω to 32.9999 Ω		1.8E-04 Ω to 6.5E-04 Ω	
	33 Ω to 109.9999 Ω		2.2E-04 Ω to 8.2E-04 Ω	
	110 Ω to 329.9999 Ω		6.4E-04 Ω to 1.2E-02 Ω	
	330 Ω to 1.09999 k Ω		1.3E-03 k Ω to 1.3E-04 k Ω	
	1.1 k Ω to 3.29999 k Ω		7.1E-03 k Ω to 1.4E-03 k Ω	
	3.3 k Ω to 10.9999 k Ω		2.5E-05 k Ω to 1.1E-04 k Ω	
	11 k Ω to 32.9999 k Ω		2.4E-03 k Ω to	

14-03-2024
Date

Sd

Director



**ACCREDITATION
DOCUMENT**

F-06/02
Issue Date: 18/08/2020
Rev. No: 09
LAB 103

Resistance (Source Mode)	33 kΩ to 109.9999 kΩ		2.5E-03 kΩ 2.5E-03 kΩ to 2.6E-03 kΩ	Fluke Warranted METCAL Procedures, Fluke 5522A Multi Product Calibrator
	110 kΩ to 329.9999 kΩ		1.0E-03 kΩ to 9.4E-03 kΩ	
	330 kΩ to 1.09999 MΩ		9.0E-03 kΩ to 1.1E-04 MΩ	
	1.1 MΩ to 3.29999 MΩ		1.3E-04 MΩ to 4.3E-04 MΩ	
	3.3 MΩ to 10.9999 MΩ		1.4E-03 MΩ to 1.5E-03 MΩ	
	11 MΩ to 32.9999 MΩ		1.9E-03 MΩ to 6.1E-03 MΩ	
	33 MΩ to 109.9999 MΩ		1.3E-01 MΩ to 1.4E-01 MΩ	
	110 MΩ to 329.9999 MΩ		1.4E-01 MΩ to 1.7E-01 MΩ	
Temperature (Source Mode)	K Type	-200 °C to -100 °C	3.6E-01°C to 3.7E-01 °C	
		-100 °C to -25 °C	1.9E-01 °C to 3.7E-01 °C	
		-25 °C to 120 °C	1.8E-01°C to 1.9E-01 °C	
		120 °C to 1000 °C	1.8E-01°C to 2.8E-01 °C	
		1000 °C to 1372 °C	2.8E-01 °C to 4.1E-01 °C	
	J Type	-210 °C to -100 °C	3.0E-01 °C to 3.2E-01 °C	
		-100 °C to -30 °C	1.9E-01 °C to 3.2E-01 °C	
		-30 °C to 150 °C	1.9E-01 °C to 2.0E-01 °C	
		150 °C to 760 °C	2.0E-01 °C to 2.1E-01 °C	
		760 °C to 1200 °C	2.0E-01 °C to 2.6E-01 °C	
	E Type	-250 °C to -100 °C	3.6E-01 °C to 3.7E-01 °C	

14-03-2024
Date

Sd
Director



**ACCREDITATION
DOCUMENT**

F-06/02
Issue Date: 18/08/2020
Rev. No: 09
LAB 103

Temperature (Source Mode)		-100 °C to -25 °C	1.9E-01 °C to 3.7E-01 °C	Fluke Warranted METCAL Procedures, Fluke 5522A Multi Product Calibrator
		-25 °C to 350 °C	1.8E-01 °C to 1.9E-01 °C	
		350 °C to 650 °C	1.8E-01 °C to 2.8E-01 °C	
		650 °C to 1000 °C	2.8E-01 °C to 4.1E-01 °C	
	N Type	-200 °C to -100 °C	4.1E-01 °C to 4.3E-01 °C	
		-100 °C to -25 °C	2.3E-01 °C to 4.3E-01 °C	
		-25 °C to 120 °C	2.2E-01 °C to 2.3E-01 °C	
		120 °C to 410 °C	2.2E-01 °C to 3.0E-01 °C	
		410 °C to 1300 °C	2.9E-01 °C to 3.0E-01 °C	
	R Type	0 °C to 250 °C	5.9E-01 °C to 6.0E-01 °C	
		250 °C to 400 °C	3.9E-01 °C to 5.9E-01 °C	
		400 °C to 1000 °C	3.6E-01 °C to 3.9E-01 °C	
		1000 °C to 1767 °C	3.6E-01 °C to 4.1E-01 °C	
	S Type	0 °C to 250 °C	5.7E-01 °C to 5.8E-01 °C	
		250 °C to 1000 °C	5.8E-01 °C to 3.7E-01 °C	
		1000 °C to 1400 °C	1.9E-01 °C to 3.7E-01 °C	
		1400 °C to 1757 °C	1.8E-01 °C to 1.9E-01 °C	
	T Type	-250 °C to -150 °C	2.8E-01 °C to 6.5E-01 °C	
		-150 °C to 0 °C	1.9E-01 °C to 2.8E-01 °C	
		0 °C to 120 °C	1.8E-01 °C to 1.9E-01 °C	
120 °C to 400 °C		1.8E-01 °C to		

14-03-2024
Date

Sd
Director

	ACCREDITATION DOCUMENT	F-06/02 Issue Date: 18/08/2020 Rev. No: 09 LAB 103
---	-----------------------------------	---

		1.9E-01 °C	
Frequency (Source Mode)	0.01 Hz to 119.99 Hz	4.8E-04 Hz to 2.2E-03 Hz	
	120 Hz to 1199.9 Hz	2.6E-03 Hz to 3.2E-03 Hz	
Frequency (Source Mode)	1.200 kHz to 11.999 kHz	1.4E-05 kHz to 5.8E-04 kHz	Fluke Warranted METCAL Procedures, Fluke 5522A Multi Product Calibrator
	12.00 kHz to 119.99 kHz	1.5E-05 kHz to 6.7E-05 kHz	
	120.0 kHz to 1199.9 kHz	6.0E-05 kHz to 1.9E-03 kHz	
	1.20 MHz	5.8E-07 MHz	

*** Expanded Uncertainty:**

- ☐ Expanded Uncertainty is the measurement uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. This measurement uncertainty is a value for which the laboratory has been accredited using the procedure that was the subject of assessment. In certificates issued under its accreditation scope an accredited laboratory is not permitted to quote an uncertainty that is smaller than the published uncertainty for respective ranges as given above.

14-03-2024
Date

Sd
Director