

	ACCREDITATION DOCUMENT	F-06/02 Issue Date: 18/08/2020 Rev. No: 09 LAB 158
---	-----------------------------------	---

Accreditation No: LAB 158

Awarded to

**AMSON MEDICAL DEVICES QUALITY CONTROL LABORATORY
AMSON VACCINES & PHARMA (PVT) Ltd.
Plot No. 113-Industrial Triangle, Kahuta Road, Islamabad, Pakistan.**

The scope of accreditation is in accordance with the standard specifications outlined in the following page(s) of this document. The accredited scope shall be visible and legible in areas such as customer service, sample-receiving section etc and shall not mislead its users.

The accreditation was first time granted on **12-06-2018** by Pakistan National Accreditation Council.

The laboratory complies with the requirements of **ISO/IEC 17025:2017**.

The accreditation requires regular surveillance, and is valid until **23-07-2024**.

The decision of accreditation made by Pakistan National Accreditation Council implies that the organization has been found to fulfill the requirements for accreditation within the scope.

The organization however, itself is responsible for the results of performed measurements/tests.

PAKISTAN NATIONAL ACCREDITATION COUNCIL

02-08-2023

Date

SD

Director General

	ACCREDITATION DOCUMENT	F-06/02 Issue Date: 18/08/2020 Rev. No: 09 LAB 158
---	-----------------------------------	---

Testing Laboratory.

Accreditation Scope of **QUALITY CONTROL LABORATORY-
AMSON VACCINES & PHARMA (PVT) Ltd.**
**Plot No. 113-Industrial Triangle, Kahuta Road, Islamabad,
Pakistan.**

Permanent laboratory premises

Materials/ Products tested	Testing field (e.g. environmental testing or mechanical testing)	Types of test/ Properties measured	Reference to standardized method (e.g. ISO 14577- 1:2003)/ Internal method reference
Medical Devices (Syringes)	Chemical Test	UV absorbance	BP 2021, Vol-5, Page # Appendix XIX G V-A647/ UV - Vis. Spectrophotometer
Medical Devices (Syringes)	Physical Test	Appearance of Solution	BP 2021, Vol-5, Page # Appendix XIX G V-A647
Medical Devices (Syringes)	Physical Test	Dead Space	ISO 7886-1:2017, Annex C

02-08-2023
Date

Sd
Director