

	<b>ACCREDITATION DOCUMENT</b>	<b>F-06/02 Issue Date: 18/08/2020 Rev. No: 09 LAB 330</b>
---	-----------------------------------	---

**Accreditation No: LAB 330**

**Awarded to**

**SNGPL CENTRAL METER SHOP  
113/14 Quaid e Azam, Industrial Estate, Kot Lakhpat  
LAHORE, PAKISTAN.**

The scope of accreditation is in accordance with the standard specifications outlined in the following page(s) of this document. The accredited scope shall be visible and legible in areas such as customer service, sample-receiving section etc and shall not mislead its users.

The accreditation was first time granted on **11-03-2025** by Pakistan National Accreditation Council.

The laboratory complies with the requirements of **ISO/IEC 17025:2017**.

The accreditation requires regular surveillance, and is valid until **10-03-2028**

The decision of accreditation made by Pakistan National Accreditation Council implies that the organization has been found to fulfill the requirements for accreditation within the scope.

The organization however, itself is responsible for the results of performed measurements/tests.

**PAKISTAN NATIONAL ACCREDITATION COUNCIL**

11-03-2025

Date

SD

Director General

	<b>ACCREDITATION DOCUMENT</b>	<b>F-06/02</b> <b>Issue Date: 18/08/2020</b> <b>Rev. No: 09</b> <b>LAB 330</b>
---	-----------------------------------	---

**Calibration Laboratory.**

Accreditation Scope of **SNGPL Central Meter Shop Lahore, Pakistan**

Permanent laboratory premises 113/14 Industrial Estate Area , Kot LakhPat  
Lahore

<b>Field of measurement:</b>			
<b>Measured quantity</b>	<b>Range</b>	<b>*Expanded Uncertainty ( ± )</b>	<b>Technique, Reference Standard, Equipment</b>
<b>TEMPERATURE CALIBRATION OF EVC</b>	60 -120 °F	60 °F = 0.014739286 80 °F = 0.01489315 120 °F = 0.014633306	Digital Temperature calibrator

11-03-2025  
Date

Sd  
Director