Accreditation No: LAB 376

Awarded to

IndTech Calibration & Inspection Services (Private) Limited.
Commercial Center-3, GECHS Phase-II,
Link Road Model Town, Lahore, Punjab, Pakistan.

The scope of accreditation is in accordance with the standard specifications outlined in the following page(s) of this document. The accredited scope shall be visible and legible in areas such as customer service, sample-receiving section etc and shall not mislead its users.

The accreditation was first time granted on **14-09-2026** by Pakistan National Accreditation Council.

The laboratory complies with the requirements of **ISO/IEC 17025:2017.**

The accreditation requires regular surveillance, and is valid until **13-09-2029.**

The decision of accreditation made by Pakistan National Accreditation Council implies that the organization has been found to fulfill the requirements for accreditation within the scope.

The organization however, itself is responsible for the results of performed measurements/tests.

PAKISTAN NATIONAL ACCREDITATION COUNCIL

14-09-2026
Date

SD
Director General

Calibration Laboratory.

Accreditation Scope of IndTech Calibration & Inspection Services (Private) Limited, Lahore, Pakistan.

Permanent Laboratory Premises ☒

Field of Measurement: Pressure Measurement and Sourcing			
Measured quantity	**Range**	***Expanded Uncertainty (±)**	**Technique, Reference Standard, Equipment**
Field of measurement: Temperature Measurement			
Temperature Source	- 20.00 ºC **to** 600.00 ºC	0.80 ºC – 1.61 ºC	**SOP: EURAMET cg-13** Digital Thermometer with probe
Temperature (Dial Gauge Thermometers)	-10.0 ºC **to** 110.0 ºC	0.69 ºC - 0.91 ºC	**SOP: ICIS/MS/SOP-004** Digital Thermometer with probe and Dry Block Calibrator.

Scope of Calibration for Onsite ☒

Measured quantity	**Range**	***Expanded Uncertainty (±)**	**Technique, Reference Standard, Equipment**
Field of measurement: Temperature Measurement			
Dial Gauge Thermometers	-10.0 ºC **to** 110.0 ºC	0.69 ºC - 0.91 ºC	**SOP: ICIS/MS/SOP-004** Digital Thermometer with probe and Dry Block Calibrator.
Temperature Bath	-20.0 °C to 95.0 ºC	0.5 ºC – 0.9 ºC	**SOP: ICIS/MS/SOP-006** Digital Thermometer with probe.

* **Expanded Uncertainty:**

- Expanded Uncertainty is the measurement uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of k = 2. This measurement uncertainty is a value for which the laboratory has been accredited using the procedure that was the subject of assessment. In certificates issued under its accreditation scope an accredited laboratory is not permitted to quote an uncertainty that is smaller than the published uncertainty for respective ranges as given above.

14-09-2026
Date

Sd
Director