

	ACCREDITATION DOCUMENT	F-06/02 Issue Date: 10/08/15 Rev. No: 07 LAB 169
---	-----------------------------------	---

Accreditation No: LAB 169

Awarded to

**Al-Mizan Industrial Solution
Azizia, Doha, Qatar P.O Box # 16690.**

The scope of accreditation is in accordance with the standard specifications outlined in the following page(s) of this document. The accredited scope shall be visible and legible in areas such as customer service, sample-receiving section etc and shall not mislead its users.

The accreditation was first time granted on **25-01-2019** by Pakistan National Accreditation Council.

The laboratory complies with the requirements of **ISO/IEC 17025:2005**.

The accreditation requires regular surveillance, and is valid until **24-01-2022**.

The decision of accreditation made by Pakistan National Accreditation Council implies that the organization has been found to fulfill the requirements for accreditation within the scope.

The organization however, itself is responsible for the results of performed measurements/tests.

PAKISTAN NATIONAL ACCREDITATION COUNCIL

25-01-2019

Date

Director General

	ACCREDITATION DOCUMENT	F-06/02 Issue Date: 10/08/15 Rev. No: 07 LAB 169
---	-----------------------------------	---

Calibration Laboratory.

Al Mizan Industrial Solutions (AMIS), Azizia, Doha Qatar.

Permanent laboratory premises

Field of measurement:			
Measured quantity	Range	Calibration & Measurement Capability (CMC) expressed as an uncertainty (\pm)	Brief description of measurement and equipment used
Pressure	-1 to +20 bar	+/- 0.015 bar	The advance modular calibrator (Druck DPI-620 with pressure modules to calibrate various types of pressure instruments like pressure gauges, pressure switches, pressure transmitters, pressure chart recorders, digital pressure indicators, pressure controllers and vacuum gauges.
Temperature	-15 to +140 °C	+/- 0.392 °C	Temperature dry Block Calibrator (Ametek-Jofra-CTC-140A) Used to calibrate Temperature Gauges, Temperature recorders, temperature switches, temperature transmitters, temperature controllers, digital temperature indicators (sensor type), thermocouples, RTDS

25-01-2019

Date

Director